

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL06 Date Analyzed: 03/31/2025

Lab Sample No.: AR1660CCC500 Data File : PP071027.D Time Analyzed: 19:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.674	5.563	5.763	532.760	500.000	6.6
Aroclor-1016-2	5.696	5.585	5.785	531.220	500.000	6.2
Aroclor-1016-3	5.758	5.648	5.848	526.790	500.000	5.4
Aroclor-1016-4	5.855	5.745	5.945	568.260	500.000	13.7
Aroclor-1016-5	6.149	6.038	6.238	588.040	500.000	17.6
Aroclor-1260-1	7.269	7.156	7.356	579.580	500.000	15.9
Aroclor-1260-2	7.523	7.410	7.610	588.530	500.000	17.7
Aroclor-1260-3	7.882	7.768	7.968	575.050	500.000	15.0
Aroclor-1260-4	8.107	7.992	8.192	546.960	500.000	9.4
Aroclor-1260-5	8.427	8.311	8.511	568.920	500.000	13.8
Decachlorobiphenyl	10.248	10.125	10.325	57.990	50.000	16.0
Tetrachloro-m-xylene	4.519	4.412	4.612	54.480	50.000	9.0

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL06 Date Analyzed: 03/31/2025

Lab Sample No.: AR1660CCC500 Data File : PP071027.D Time Analyzed: 19:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.905	4.803	5.003	533.670	500.000	6.7
Aroclor-1016-2	4.923	4.822	5.022	523.190	500.000	4.6
Aroclor-1016-3	5.101	4.999	5.199	555.200	500.000	11.0
Aroclor-1016-4	5.142	5.041	5.241	570.400	500.000	14.1
Aroclor-1016-5	5.358	5.255	5.455	564.150	500.000	12.8
Aroclor-1260-1	6.394	6.292	6.492	563.190	500.000	12.6
Aroclor-1260-2	6.582	6.480	6.680	551.480	500.000	10.3
Aroclor-1260-3	6.736	6.634	6.834	583.150	500.000	16.6
Aroclor-1260-4	7.208	7.105	7.305	542.190	500.000	8.4
Aroclor-1260-5	7.449	7.346	7.546	498.080	500.000	-0.4
Decachlorobiphenyl	8.864	8.759	8.959	51.370	50.000	2.7
Tetrachloro-m-xylene	3.817	3.716	3.916	52.750	50.000	5.5