

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1664

Client: Weston Solutions, Inc.

Analytical Method: 8081B

DataFile : PL094965.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	P001-BBDGA-001-01MS											
Q1664-05MS	alpha-BHC	0.5	0	0.55	ug/L	110				46	160	
	beta-BHC	0.5	0	0.55	ug/L	111				15	175	
	delta-BHC	0.5	0	0.55	ug/L	110				10	175	
	gamma-BHC (Lindane)	0.5	0	0.55	ug/L	110				60	152	
	Heptachlor	0.5	0	0.48	ug/L	95				56	147	
	Aldrin	0.5	0	0.47	ug/L	94				45	147	
	Heptachlor epoxide	0.5	0	0.53	ug/L	106				77	143	
	Endosulfan I	0.5	0	0.53	ug/L	107				34	157	
	Dieldrin	0.5	0	0.54	ug/L	108				46	155	
	4,4'-DDE	0.5	0	0.55	ug/L	109				36	162	
	Endrin	0.5	0	0.56	ug/L	111				76	144	
	Endosulfan II	0.5	0	0.54	ug/L	109				21	168	
	4,4'-DDD	0.5	0	0.60	ug/L	119				15	175	
	Endosulfan sulfate	0.5	0	0.55	ug/L	111				14	183	
	4,4'-DDT	0.5	0	0.56	ug/L	112				15	175	
	Methoxychlor	0.5	0	0.58	ug/L	116				70	142	
	Endrin ketone	0.5	0	0.56	ug/L	111				25	172	
	Endrin aldehyde	0.5	0	0.51	ug/L	102				28	175	
	alpha-Chlordane	0.5	0	0.52	ug/L	104				34	160	
	gamma-Chlordane	0.5	0	0.52	ug/L	104				31	163	