

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1664

Client: Weston Solutions, Inc.

Analytical Method: 8081B

DataFile : PL094966.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	P001-BBDGA-001-01MSD											
Q1664-06MSD	alpha-BHC	0.5	0	0.52	ug/L	105		5		46	160	20
	beta-BHC	0.5	0	0.53	ug/L	105		6		15	175	20
	delta-BHC	0.5	0	0.52	ug/L	104		6		10	175	20
	gamma-BHC (Lindane)	0.5	0	0.53	ug/L	105		5		60	152	20
	Heptachlor	0.5	0	0.45	ug/L	90		5		56	147	20
	Aldrin	0.5	0	0.44	ug/L	87		8		45	147	22
	Heptachlor epoxide	0.5	0	0.50	ug/L	100		6		77	143	20
	Endosulfan I	0.5	0	0.51	ug/L	101		6		34	157	20
	Dieldrin	0.5	0	0.51	ug/L	101		7		46	155	20
	4,4'-DDE	0.5	0	0.52	ug/L	104		5		36	162	20
	Endrin	0.5	0	0.52	ug/L	105		6		76	144	20
	Endosulfan II	0.5	0	0.51	ug/L	103		6		21	168	20
	4,4'-DDD	0.5	0	0.57	ug/L	113		5		15	175	20
	Endosulfan sulfate	0.5	0	0.52	ug/L	104		7		14	183	20
	4,4'-DDT	0.5	0	0.53	ug/L	106		6		15	175	20
	Methoxychlor	0.5	0	0.54	ug/L	109		6		70	142	20
	Endrin ketone	0.5	0	0.52	ug/L	105		6		25	172	20
	Endrin aldehyde	0.5	0	0.49	ug/L	97		5		28	175	20
	alpha-Chlordane	0.5	0	0.49	ug/L	98		6		34	160	20
	gamma-Chlordane	0.5	0	0.49	ug/L	97		7		31	163	20