

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1664

Client: Weston Solutions, Inc.

Analytical Method: **8081B** Datafile : PL095030.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167395BS	alpha-BHC	0.5	0.49	ug/L	98				85	130	
	beta-BHC	0.5	0.47	ug/L	94				83	126	
	delta-BHC	0.5	0.46	ug/L	92				69	141	
	gamma-BHC (Lindane)	0.5	0.47	ug/L	94				82	129	
	Heptachlor	0.5	0.45	ug/L	90				79	127	
	Aldrin	0.5	0.45	ug/L	90				79	126	
	Heptachlor epoxide	0.5	0.46	ug/L	92				81	124	
	Endosulfan I	0.5	0.47	ug/L	93				85	122	
	Dieldrin	0.5	0.48	ug/L	96				83	125	
	4,4'-DDE	0.5	0.51	ug/L	103				80	127	
	Endrin	0.5	0.48	ug/L	96				81	128	
	Endosulfan II	0.5	0.48	ug/L	96				82	123	
	4,4'-DDD	0.5	0.55	ug/L	110				77	131	
	Endosulfan sulfate	0.5	0.49	ug/L	98				76	129	
	4,4'-DDT	0.5	0.51	ug/L	101				80	133	
	Methoxychlor	0.5	0.53	ug/L	107				78	108	
	Endrin ketone	0.5	0.52	ug/L	104				80	131	
	Endrin aldehyde	0.5	0.48	ug/L	96				82	127	
	alpha-Chlordane	0.5	0.46	ug/L	92				82	125	
	gamma-Chlordane	0.5	0.46	ug/L	92				82	125	