

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/03/2025

Lab File ID: BG064164.D Time Analyzed: 13:04

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	32960	7.856	151982	10.65	113011	14.48
UPPER LIMIT	65920	8.356	303964	11.152	226022	14.983
LOWER LIMIT	16480	7.356	75991	10.152	56505.5	13.983
EPA SAMPLE NO.						
01 PB167393BL	25692	7.86	118561	10.65	97336	14.48
02 PB167393BS	28326	7.86	137392	10.65	107393	14.48
03 PB167393TB	29004	7.86	129149	10.65	102109	14.49

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.