

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF031025

Review By	anahy	Review On	3/11/2025 9:11:47 AM
Supervise By	Jagrut	Supervise On	3/11/2025 10:18:40 AM
SubDirectory	BF031025	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141896.D	10 Mar 2025 10:31		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF141897.D	10 Mar 2025 11:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF141898.D	10 Mar 2025 11:30	Compound#9,32,54,65 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF141899.D	10 Mar 2025 12:00		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF141900.D	10 Mar 2025 12:29	Compound #54 Kept on LR, Method is good for DOD . Method failed for compound #77.	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF141901.D	10 Mar 2025 12:58		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF141902.D	10 Mar 2025 13:28	Not used	RC/JU	Not Ok
8	SSTDICC060	SSTDICC060	BF141903.D	10 Mar 2025 13:57	Compound#69 Removed from 60 ppm	RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF141904.D	10 Mar 2025 14:27	Compound#9,69 Removed from 80 ppm	RC/JU	Ok,M
10	SSTDICC050	SSTDICC050	BF141905.D	10 Mar 2025 15:20		RC/JU	Ok
11	SSTDICV040	ICVBF031025	BF141906.D	10 Mar 2025 15:53		RC/JU	Ok
12	PB167012BL	PB167012BL	BF141907.D	10 Mar 2025 17:09		RC/JU	Ok
13	SP6752	SP6752	BF141908.D	10 Mar 2025 17:39	8270 SPIKE-SP6752	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF033125

Review By	anahy	Review On	4/1/2025 11:09:34 AM
Supervise By	Jagrut	Supervise On	4/1/2025 3:40:46 PM
SubDirectory	BF033125	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12656,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142174.D	31 Mar 2025 11:53		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142175.D	31 Mar 2025 12:22		RC/JU	Ok,M
3	PB167369BL	PB167369BL	BF142176.D	31 Mar 2025 12:52		RC/JU	Ok
4	PB167369BS	PB167369BS	BF142177.D	31 Mar 2025 13:22		RC/JU	Ok,M
5	PB167233TB	PB167233TB	BF142178.D	31 Mar 2025 13:51		RC/JU	Ok,M
6	Q1657-01DL	72-11998DL	BF142179.D	31 Mar 2025 14:29		RC/JU	Ok,M
7	Q1650-01DL	STOCK-PILE-BIN2DL	BF142180.D	31 Mar 2025 14:59	Internal Standard Fail	RC/JU	Not Ok
8	Q1654-01DL	RT3407DL	BF142181.D	31 Mar 2025 15:28	Internal Standard Fail, Need 10X further Dilution	RC/JU	Not Ok
9	Q1664-09	P001-BBDGA-002-01	BF142182.D	31 Mar 2025 15:58	Internal Standard Fail	RC/JU	ReRun
10	Q1654-01DL2	RT3407DL2	BF142183.D	31 Mar 2025 16:28	Internal Standard Fail	RC/JU	Not Ok
11	Q1664-11	P001-BBDGA-003-01	BF142184.D	31 Mar 2025 16:58		RC/JU	Ok,M
12	Q1664-13	P001-BBDGA-004-01	BF142185.D	31 Mar 2025 17:27		RC/JU	Ok,M
13	Q1664-17	P001-BBDGA-006-01	BF142186.D	31 Mar 2025 17:57	Internal Standard Fail	RC/JU	ReRun
14	Q1664-21	P001-BBDGA-008-01	BF142187.D	31 Mar 2025 18:27		RC/JU	Ok,M
15	Q1664-01	P001-BBDGA-001-01	BF142188.D	31 Mar 2025 18:57		RC/JU	Ok,M
16	Q1664-15	P001-BBDGA-005-01	BF142189.D	31 Mar 2025 19:26	Internal Standard Fail	RC/JU	ReRun
17	Q1664-19	P001-BBDGA-007-01	BF142190.D	31 Mar 2025 19:56		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF033125

Review By	anahy	Review On	4/1/2025 11:09:34 AM		
Supervise By	Jagrut	Supervise On	4/1/2025 3:40:46 PM		
SubDirectory	BF033125	HP Acquire Method	BNA_F	HP Processing Method	bf031025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729				
CCC	SP6725				
Internal Standard/PEM	S12656,10ul/1000ul sample				
ICV/I.BLK	SP6686				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1664-07	P001-BBDGA-001-02	BF142191.D	31 Mar 2025 20:26	Internal Standard Fail	RC/JU	ReRun
19	Q1664-02MS	P001-BBDGA-001-01M	BF142192.D	31 Mar 2025 20:55	Internal Standard Fail	RC/JU	Not Ok
20	Q1664-03MSD	P001-BBDGA-001-01M	BF142193.D	31 Mar 2025 21:26	Internal Standard Fail	RC/JU	Not Ok
21	Q1671-01	WC-1	BF142194.D	31 Mar 2025 21:55	Internal Standard Fail	RC/JU	ReRun
22	Q1674-01	RT5358	BF142195.D	31 Mar 2025 22:25	Internal Standard Fail	RC/JU	ReRun
23	Q1671-04	WC-1	BF142196.D	31 Mar 2025 22:55	Internal Standard Fail	RC/JU	ReRun
24	Q1674-03	72-11991	BF142197.D	31 Mar 2025 23:25	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM031325

Review By	anahy	Review On	3/14/2025 4:40:45 PM
Supervise By	Jagrut	Supervise On	3/14/2025 4:41:58 PM
SubDirectory	BM031325	HP Acquire Method	BNA_M
		HP Processing Method	bm031325

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM049705.D	13 Mar 2025 08:24		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM049706.D	13 Mar 2025 09:02		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM049707.D	13 Mar 2025 09:42	Compound#9,32,54,56,65,70,77, 82,85 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM049708.D	13 Mar 2025 10:21		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM049709.D	13 Mar 2025 11:00		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BM049710.D	13 Mar 2025 11:39	Compound#62 Kept on LR & 54,65 Kept on QR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BM049711.D	13 Mar 2025 12:19		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM049712.D	13 Mar 2025 12:58	Compound#69 removed from 60 ppm	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BM049713.D	13 Mar 2025 13:37	Compound#9,42,69 removed from 80 ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBM031325	BM049714.D	13 Mar 2025 14:41		RC/JU	Ok,M
11	PB167097BL	PB167097BL	BM049715.D	13 Mar 2025 15:20	Use for DOD Projects only	RC/JU	Ok
12	Q1502-22	RR-TRIAZINE-WP	BM049716.D	13 Mar 2025 16:00	PT Sample - Already analyzed.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM040225

Review By	anahy	Review On	4/3/2025 11:57:14 AM		
Supervise By	Jagrut	Supervise On	4/3/2025 4:20:55 PM		
SubDirectory	BM040225	HP Acquire Method	BNA_M	HP Processing Method	bm031325
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12657,10ul/1000ul sample			
ICV/I.BLK		SP6686			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM049783.D	02 Apr 2025 09:29		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM049784.D	02 Apr 2025 10:46		RC/JU	Ok
3	PB167406BL	PB167406BL	BM049785.D	02 Apr 2025 11:25	Surrogate fail	RC/JU	Not Ok
4	PB167406BS	PB167406BS	BM049786.D	02 Apr 2025 12:04		RC/JU	Ok,M
5	Q1679-01	TP-10	BM049787.D	02 Apr 2025 12:49		RC/JU	Ok
6	Q1679-05	TP-9	BM049788.D	02 Apr 2025 13:28		RC/JU	Ok,M
7	Q1680-01	TP-4	BM049789.D	02 Apr 2025 14:07		RC/JU	Ok
8	Q1680-01MS	TP-4MS	BM049790.D	02 Apr 2025 14:46		RC/JU	Ok,M
9	Q1680-01MSD	TP-4MSD	BM049791.D	02 Apr 2025 15:25		RC/JU	Ok,M
10	Q1681-01	TP-12	BM049792.D	02 Apr 2025 16:04		RC/JU	Ok
11	Q1664-09RE	P001-BBDGA-002-01R	BM049793.D	02 Apr 2025 16:43	Fax is already given	RC/JU	Confirms
12	Q1664-17RE	P001-BBDGA-006-01R	BM049794.D	02 Apr 2025 17:22	Fax is already given	RC/JU	Confirms
13	Q1664-15RE	P001-BBDGA-005-01R	BM049795.D	02 Apr 2025 18:00	Fax is already given	RC/JU	Confirms
14	Q1664-07RE	P001-BBDGA-001-02R	BM049796.D	02 Apr 2025 18:39	Fax is already given	RC/JU	Confirms
15	Q1664-02MS	P001-BBDGA-001-01M	BM049797.D	02 Apr 2025 19:18		RC/JU	Ok
16	Q1664-03MSD	P001-BBDGA-001-01M	BM049798.D	02 Apr 2025 19:57		RC/JU	Ok

M : Manual Integration