

Surrogate Summary

SDG No.: **Q1664**

Client: **Weston Solutions, Inc.**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1664-04	P001-BBDGA-001-01	1,2-Dichloroethane-d4	50	57.4	115	74	125
		Dibromofluoromethane	50	53.5	107	75	124
		Toluene-d8	50	49.0	98	86	113
		4-Bromofluorobenzene	50	44.8	90	77	121
Q1664-05MS	P001-BBDGA-001-01MS	1,2-Dichloroethane-d4	50	43.8	88	74	125
		Dibromofluoromethane	50	39.0	78	75	124
		Toluene-d8	50	48.1	96	86	113
		4-Bromofluorobenzene	50	50.6	101	77	121
Q1664-06MSD	P001-BBDGA-001-01MSD	1,2-Dichloroethane-d4	50	48.4	97	74	125
		Dibromofluoromethane	50	41.4	83	75	124
		Toluene-d8	50	51.8	104	86	113
		4-Bromofluorobenzene	50	52.6	105	77	121
Q1664-08	P001-BBDGA-001-02	1,2-Dichloroethane-d4	50	57.8	116	74	125
		Dibromofluoromethane	50	52.3	105	75	124
		Toluene-d8	50	47.2	94	86	113
		4-Bromofluorobenzene	50	42.6	85	77	121
Q1664-10	P001-BBDGA-002-01	1,2-Dichloroethane-d4	50	56.8	114	74	125
		Dibromofluoromethane	50	53.4	107	75	124
		Toluene-d8	50	48.4	97	86	113
		4-Bromofluorobenzene	50	44.3	89	77	121
Q1664-12	P001-BBDGA-003-01	1,2-Dichloroethane-d4	50	56.5	113	74	125
		Dibromofluoromethane	50	52.2	104	75	124
		Toluene-d8	50	48.0	96	86	113
		4-Bromofluorobenzene	50	44.0	88	77	121
Q1664-14	P001-BBDGA-004-01	1,2-Dichloroethane-d4	50	57.6	115	74	125
		Dibromofluoromethane	50	53.5	107	75	124
		Toluene-d8	50	48.6	97	86	113
		4-Bromofluorobenzene	50	45.5	91	77	121
Q1664-16	P001-BBDGA-005-01	1,2-Dichloroethane-d4	50	57.3	115	74	125
		Dibromofluoromethane	50	52.8	106	75	124
		Toluene-d8	50	49.0	98	86	113
		4-Bromofluorobenzene	50	45.4	91	77	121
Q1664-18	P001-BBDGA-006-01	1,2-Dichloroethane-d4	50	59.0	118	74	125
		Dibromofluoromethane	50	52.9	106	75	124
		Toluene-d8	50	48.1	96	86	113
		4-Bromofluorobenzene	50	45.0	90	77	121
Q1664-20	P001-BBDGA-007-01	1,2-Dichloroethane-d4	50	59.2	118	74	125
		Dibromofluoromethane	50	54.1	108	75	124
		Toluene-d8	50	49.2	98	86	113
		4-Bromofluorobenzene	50	47.0	94	77	121
Q1664-22	P001-BBDGA-008-01	1,2-Dichloroethane-d4	50	58.7	117	74	125
		Dibromofluoromethane	50	54.5	109	75	124
		Toluene-d8	50	49.5	99	86	113
		4-Bromofluorobenzene	50	45.8	92	77	121

Surrogate Summary

SDG No.: Q1664

Client: Weston Solutions, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN0328WBL01	VN0328WBL01	1,2-Dichloroethane-d4	50	57.5	115	74	125
		Dibromofluoromethane	50	54.0	108	75	124
		Toluene-d8	50	48.0	96	86	113
		4-Bromofluorobenzene	50	41.9	84	77	121
VN0328WBS01	VN0328WBS01	1,2-Dichloroethane-d4	50	51.0	102	74	125
		Dibromofluoromethane	50	52.6	105	75	124
		Toluene-d8	50	54.8	110	86	113
		4-Bromofluorobenzene	50	51.7	103	77	121