

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ROYF02
 Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664
 Lab File ID: VN086140.D Date Analyzed: 03/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:44
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	177687	13.788				
UPPER LIMIT	355374	14.288				
LOWER LIMIT	88843.5	13.288				
EPA SAMPLE NO.						
P001-BBDGA-001-01	115525	13.79				
P001-BBDGA-001-01MS	170825	13.79				
P001-BBDGA-001-01MSD	165361	13.79				
P001-BBDGA-001-02	114450	13.79				
P001-BBDGA-002-01	126117	13.79				
P001-BBDGA-003-01	129139	13.79				
P001-BBDGA-004-01	129697	13.79				
P001-BBDGA-005-01	131137	13.79				
P001-BBDGA-006-01	126533	13.79				
P001-BBDGA-007-01	130900	13.79				
P001-BBDGA-008-01	130901	13.79				
VN0328WBL01	111030	13.79				
VN0328WBS01	178334	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.