

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM				
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM				
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m		
STD. NAME		STD REF.#					
Tune/Reschk		VP133499					
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508					
CCC		VP133510					
Internal Standard/PEM		VP131783					
ICV/I.BLK		VP133509					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021655.D	27 Mar 2025 09:23		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021656.D	27 Mar 2025 10:08		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY021657.D	27 Mar 2025 10:31		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021658.D	27 Mar 2025 10:54	Comp.#20 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021659.D	27 Mar 2025 11:16		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021660.D	27 Mar 2025 11:39		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021661.D	27 Mar 2025 12:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021662.D	27 Mar 2025 12:35		SY/MD	Ok
9	VSTDICV050	ICVVY032725	VY021663.D	27 Mar 2025 13:15		SY/MD	Ok,M
10	VY0327SBL01	VY0327SBL01	VY021664.D	27 Mar 2025 13:59		SY/MD	Ok
11	VY0327SBS01	VY0327SBS01	VY021665.D	27 Mar 2025 14:38		SY/MD	Ok,M
12	VY0327SBSD01	VY0327SBSD01	VY021666.D	27 Mar 2025 15:00		SY/MD	Ok,M
13	Q1661-02	351	VY021667.D	27 Mar 2025 15:24	vial-B Not purge	SY/MD	Not Ok
14	Q1645-03RE	TP-4-VOCRE	VY021668.D	27 Mar 2025 15:48	vial-B Internal Standard Fail	SY/MD	Confirms
15	Q1662-03	TP-6-VOC	VY021669.D	27 Mar 2025 16:11	vial-A NOT PURGE	SY/MD	Not Ok
16	Q1662-07	TP-7-VOC	VY021670.D	27 Mar 2025 16:35	vial-A Internal Standard Fail	SY/MD	ReRun
17	Q1661-02	351	VY021671.D	27 Mar 2025 16:58	vial-A NOT PURGE	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM		
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133499			
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508			
CCC		VP133510			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133509			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1650-02	STOCK-PILE-BIN2	VY021672.D	27 Mar 2025 17:21	vial-A NOT PURGE	SY/MD	Not Ok
19	Q1654-01	RT3407	VY021673.D	27 Mar 2025 17:45	vial-A Internal Standard Fail	SY/MD	Ok
20	Q1656-01	VNJ239	VY021674.D	27 Mar 2025 18:08	vial-A NOT PURGE	SY/MD	Not Ok
21	Q1648-09	TP-7	VY021675.D	27 Mar 2025 18:32	vial-A	SY/MD	Ok
22	Q1648-07	TP-6	VY021676.D	27 Mar 2025 18:55	vial-A	SY/MD	Ok
23	Q1648-05	TP-3	VY021677.D	27 Mar 2025 19:19	vial-A	SY/MD	Ok
24	Q1648-01	TP-4	VY021678.D	27 Mar 2025 19:42	vial-A	SY/MD	Ok
25	Q1648-03	TP-5	VY021679.D	27 Mar 2025 20:05	vial-A	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY021680.D	27 Mar 2025 20:28		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032825

Review By	Maresh Dadoda	Review On	3/31/2025 9:56:41 AM
Supervise By	Semsettin Yesilyurt	Supervise On	3/31/2025 9:57:25 AM
SubDirectory	VY032825	HP Acquire Method	MSVOA_Y HP Processing Method 82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133522
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133520,VP133521 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021681.D	28 Mar 2025 10:11		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021682.D	28 Mar 2025 10:42		SY/MD	Ok,M
3	VY0328SBL01	VY0328SBL01	VY021683.D	28 Mar 2025 11:21		SY/MD	Ok
4	VY0328SBS01	VY0328SBS01	VY021684.D	28 Mar 2025 11:55		SY/MD	Ok,M
5	VY0328SBSD01	VY0328SBSD01	VY021685.D	28 Mar 2025 12:24		SY/MD	Ok,M
6	Q1662-07	TP-7-VOC	VY021686.D	28 Mar 2025 12:54	vial-B	SY/MD	Ok
7	Q1656-01	VNJ239	VY021687.D	28 Mar 2025 13:17	vial-B ISTD Fail	SY/MD	Ok
8	Q1637-03	EO-02-032525	VY021688.D	28 Mar 2025 13:41	vial-B NOT PURGE	SY/MD	Not Ok
9	IBLK	IBLK	VY021689.D	28 Mar 2025 14:04	NOT PURGE	SY/MD	Not Ok
10	Q1648-11	TP-8	VY021690.D	28 Mar 2025 14:28	vial-A	SY/MD	Ok
11	Q1648-13	TP-2	VY021691.D	28 Mar 2025 14:51	vial-A	SY/MD	Ok
12	Q1650-02	STOCK-PILE-BIN2	VY021692.D	28 Mar 2025 15:14	vial-B	SY/MD	Ok
13	Q1654-01	RT3407	VY021693.D	28 Mar 2025 15:38	vial-B NOT PURGE	SY/MD	Not Ok
14	Q1662-03	TP-6-VOC	VY021694.D	28 Mar 2025 16:01	vial-B	SY/MD	Ok
15	Q1664-01	P001-BBDGA-001-01	VY021695.D	28 Mar 2025 16:25	vial-A	SY/MD	Ok
16	Q1664-02MS	P001-BBDGA-001-01M	VY021696.D	28 Mar 2025 16:48	vial-A Not spiked	SY/MD	Not Ok
17	Q1664-03MSD	P001-BBDGA-001-01M	VY021697.D	28 Mar 2025 17:12	vial-A Not spiked; added MSD weight in logbook please check	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032825

Review By	Mahesh Dadoda	Review On	3/31/2025 9:56:41 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/31/2025 9:57:25 AM		
SubDirectory	VY032825	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133522			
CCC		VP133520,VP133521			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1664-07	P001-BBDGA-001-02	VY021698.D	28 Mar 2025 17:35	vial-A	SY/MD	Ok
19	Q1664-09	P001-BBDGA-002-01	VY021699.D	28 Mar 2025 17:58	vial-A	SY/MD	Ok
20	Q1664-11	P001-BBDGA-003-01	VY021700.D	28 Mar 2025 18:22	vial-A	SY/MD	Ok
21	Q1664-13	P001-BBDGA-004-01	VY021701.D	28 Mar 2025 18:45	vial-A	SY/MD	Ok
22	Q1664-15	P001-BBDGA-005-01	VY021702.D	28 Mar 2025 19:09	vial-A	SY/MD	Ok
23	Q1664-17	P001-BBDGA-006-01	VY021703.D	28 Mar 2025 19:32	vial-A	SY/MD	Ok
24	Q1664-19	P001-BBDGA-007-01	VY021704.D	28 Mar 2025 19:55	vial-A	SY/MD	Ok
25	Q1664-21	P001-BBDGA-008-01	VY021705.D	28 Mar 2025 20:19	vial-A	SY/MD	Ok
26	Q1643-01	030825-A	VY021706.D	28 Mar 2025 20:42	vial-A	SY/MD	Ok
27	Q1657-01	72-11998	VY021707.D	28 Mar 2025 21:06	vial-A Internal Standard Fail	SY/MD	ReRun
28	VSTDCCC050	VSTDCCC050EC	VY021708.D	28 Mar 2025 21:28		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY040125

Review By	Amit Patel	Review On	4/2/2025 2:13:19 PM
Supervise By	Mahesh Dadoda	Supervise On	4/2/2025 2:15:14 PM
SubDirectory	VY040125	HP Acquire Method	MSVOA_Y HP Processing Method 82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133555
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133556,VP133557 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021730.D	01 Apr 2025 09:35		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021731.D	01 Apr 2025 10:13		SY/MD	Ok,M
3	VY0401SBL01	VY0401SBL01	VY021732.D	01 Apr 2025 10:52		SY/MD	Ok
4	VY0401SBS01	VY0401SBS01	VY021733.D	01 Apr 2025 11:30		SY/MD	Ok,M
5	VY0401SBSD01	VY0401SBSD01	VY021734.D	01 Apr 2025 11:52		SY/MD	Ok,M
6	Q1681-03	TP-12-VOC	VY021735.D	01 Apr 2025 12:36	vial-A	SY/MD	Ok
7	Q1680-03	TP-4-VOC	VY021736.D	01 Apr 2025 12:59	vial-A	SY/MD	Ok
8	Q1675-09	T3	VY021737.D	01 Apr 2025 13:22	vial-B	SY/MD	Ok
9	Q1675-12	T6	VY021738.D	01 Apr 2025 13:46	vial-A	SY/MD	Ok
10	IBLK	IBLK	VY021739.D	01 Apr 2025 14:19		SY/MD	Ok
11	Q1664-02MS	P001-BBDGA-001-01M	VY021740.D	01 Apr 2025 14:42	vial-B Internal Standard Fail	SY/MD	Ok,M
12	Q1664-03MSD	P001-BBDGA-001-01M	VY021741.D	01 Apr 2025 15:05	vial-B	SY/MD	Ok,M
13	IBLK	IBLK	VY021742.D	01 Apr 2025 15:28		SY/MD	Ok
14	Q1687-01	72-12016	VY021743.D	01 Apr 2025 15:52	Not purged	SY/MD	Not Ok
15	Q1688-01	OR-02-040125	VY021744.D	01 Apr 2025 16:15	vial-A Not purge	SY/MD	Not Ok
16	Q1692-01	NB-08-04012025	VY021745.D	01 Apr 2025 16:38	vial-A Internal Standard Fail	SY/MD	ReRun
17	VSTDCCC050	VSTDCCC050EC	VY021746.D	01 Apr 2025 17:01		SY/MD	Ok,M

M : Manual Integration