

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ROYF02
 Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664
 Lab File ID: VY021682.D Date Analyzed: 03/28/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:42
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	151090	13.346				
UPPER LIMIT	302180	13.846				
LOWER LIMIT	75545	12.846				
EPA SAMPLE NO.						
P001-BBDGA-001-01	152172	13.35				
P001-BBDGA-001-02	153193	13.35				
P001-BBDGA-002-01	154270	13.35				
P001-BBDGA-003-01	150733	13.35				
P001-BBDGA-004-01	162086	13.35				
P001-BBDGA-005-01	155440	13.35				
P001-BBDGA-006-01	151297	13.35				
P001-BBDGA-007-01	142833	13.35				
P001-BBDGA-008-01	149200	13.35				
VY0328SBL01	144468	13.35				
VY0328SBS01	151928	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.