

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ROYF02
 Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664
 Lab File ID: VY021731.D Date Analyzed: 04/01/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	185609	13.346				
UPPER LIMIT	371218	13.846				
LOWER LIMIT	92804.5	12.846				
EPA SAMPLE NO.						
P001-BBDGA-001-01MS	79914 *	13.35				
P001-BBDGA-001-01MSD	151305	13.35				
VY0401SBL01	199303	13.35				
VY0401SBS01	179014	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.