

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL022525

Review By	MMDadoda	Review On	2/26/2025 1:45:21 PM		
Supervise By		Supervise On			
SubDirectory	VL022525	HP Acquire Method	TO15AIR.M	HP Processing Method	VL022525AIR.M
STD. NAME	STD REF.#				
Tune/Reschk	AP2576				
Initial Calibration Stds	AP2582,AP2584,AP2585				
CCC	AP2582				
Internal Standard/PEM	AP2576				
ICV/I.BLK	AP2586				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042040.D	25 Feb 2025 07:56	SY/MD	Ok
2	VSTDICCC010	VL042041.D	25 Feb 2025 08:56	SY/MD	Ok,M
3	VSTDICCC002	VL042042.D	25 Feb 2025 09:29	SY/MD	Ok,M
4	VSTDICCC001	VL042043.D	25 Feb 2025 10:00	SY/MD	Ok,M
5	VSTDICCC0.5	VL042044.D	25 Feb 2025 10:31	SY/MD	Ok,M
6	VSTDICCC0.1	VL042045.D	25 Feb 2025 11:03	SY/MD	Ok,M
7	VSTDICCC0.03	VL042046.D	25 Feb 2025 11:34	SY/MD	Ok,M
8	VSTDICCC015	VL042047.D	25 Feb 2025 12:07	SY/MD	Ok,M
9	VSTDICV010	VL042048.D	25 Feb 2025 13:15	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL022625

Review By	MMDadoda	Review On	2/27/2025 12:13:24 PM
Supervise By		Supervise On	
SubDirectory	VL022625	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL022525AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2576		
Initial Calibration Stds	AP2582,AP2584,AP2585		
CCC	AP2582		
Internal Standard/PEM	AP2576		
ICV/I.BLK	AP2586		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL042049.D	26 Feb 2025 07:46	SY/MD	Ok
2	VSTDCCC010	VL042050.D	26 Feb 2025 08:38	SY/MD	Ok,M
3	VL0226ABL01	VL042051.D	26 Feb 2025 09:20	SY/MD	Ok
4	QC021725 CAN 10271	VL042052.D	26 Feb 2025 10:04	SY/MD	Ok
5	QC022525 CAN 10795	VL042053.D	26 Feb 2025 10:36	SY/MD	Ok
6	VL0226ABS01	VL042054.D	26 Feb 2025 11:07	SY/MD	Ok,M
7	Q1419-02	VL042055.D	26 Feb 2025 11:40	SY/MD	Ok,M
8	Q1430-05	VL042056.D	26 Feb 2025 12:14	SY/MD	Ok,M
9	Q1430-05DUP	VL042057.D	26 Feb 2025 12:48	SY/MD	Ok,M
10	Q1430-06	VL042058.D	26 Feb 2025 13:21	SY/MD	Ok,M
11	Q1430-06DL	VL042059.D	26 Feb 2025 13:52	SY/MD	Not Ok
12	Q1419-01	VL042060.D	26 Feb 2025 14:24	SY/MD	Dilution
13	Q1419-01DL	VL042061.D	26 Feb 2025 14:55	SY/MD	Not Ok
14	Q1419-01DL	VL042062.D	26 Feb 2025 15:27	SY/MD	Ok,M
15	Q1419-01DL2	VL042063.D	26 Feb 2025 15:58	SY/MD	Not Ok
16	VSTDCCC010	VL042064.D	26 Feb 2025 16:29	SY/MD	Ok,M

M : Manual Integration

Process Report

Order ID : Q1665

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
03/31/2025	B2502051	Q1665-19	-8.0	10331	6 L	10206		C2502003	VL042078.D Seq :VL030325 TUNE : VL042074.D ICAL : VL042046.D CCAL : VL042075.D MB : VL042076.D
03/31/2025	B2502051	Q1665-21	-7.5	10332	6 L	10236		C2502002	VL042052.D Seq :VL022625 TUNE : VL042049.D ICAL : VL042046.D CCAL : VL042050.D MB : VL042051.D
03/31/2025	B2502051	Q1665-16	-8.0	10608	6 L	10240		C2502001	VL041974.D Seq :VL020325 TUNE : VL041957.D ICAL : VL041894.D CCAL : VL041958.D MB : VL041960.D
03/31/2025	B2502051	Q1665-09	-5.5	10599	6 L	10207		C2502001	VL041974.D Seq :VL020325 TUNE : VL041957.D ICAL : VL041894.D CCAL : VL041958.D MB : VL041960.D
03/31/2025	B2502051	Q1665-11	-6.3	10598	6 L	10515		C2502001	VL041974.D Seq :VL020325 TUNE : VL041957.D ICAL : VL041894.D CCAL : VL041958.D MB : VL041960.D
03/31/2025	B2502051	Q1665-20	-5.5	10595	6 L	10626		C2502002	VL042052.D Seq :VL022625 TUNE : VL042049.D ICAL : VL042046.D CCAL : VL042050.D MB : VL042051.D

Process Report

Order ID : Q1665

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
03/31/2025	B2502051	Q1665-07	-6.7	10592	6 L	10558		C2502002	VL042052.D Seq :vl022625 TUNE : VL042049.D ICAL : VL042046.D CCAL : VL042050.D MB : VL042051.D
03/31/2025	B2502051	Q1665-10	-7.1	10492	6 L	10172		C2503001	VL042143.D Seq :VL031425 TUNE : VL042140.D ICAL : VL042046.D CCAL : VL042141.D MB : VL042142.D
03/31/2025	B2502051	Q1665-02	-3.5	10443	6 L	10203		C2501001	VL041853.D Seq :vl010725 TUNE : VL041836.D ICAL : VL041843.D CCAL : VL041838.D MB : VL041846.D
03/31/2025	B2502051	Q1665-01	-8.8	10323	6 L	10478		C2502001	VL041974.D Seq :vl020325 TUNE : VL041957.D ICAL : VL041894.D CCAL : VL041958.D MB : VL041960.D
03/31/2025	B2502051	Q1665-13	-1.8	10315	6 L	10183		C2502003	VL042078.D Seq :VL030325 TUNE : VL042074.D ICAL : VL042046.D CCAL : VL042075.D MB : VL042076.D
03/31/2025	B2502051	Q1665-22	-7.1	10306	6 L	10713		C2502002	VL042052.D Seq :vl022625 TUNE : VL042049.D ICAL : VL042046.D CCAL : VL042050.D MB : VL042051.D
03/31/2025	B2502051	Q1665-18	-8.0	10297	6 L	10190		C2502001	VL041974.D Seq :vl020325 TUNE : VL041957.D ICAL : VL041894.D CCAL : VL041958.D MB : VL041960.D

Process Report

Order ID : Q1665

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
03/31/2025	B2502051	Q1665-17	-6.7	10294	6 L	10505		C2502003	VL042078.D Seq :VL030325 TUNE : VL042074.D ICAL : VL042046.D CCAL : VL042075.D MB : VL042076.D
03/31/2025	B2502051	Q1665-05	-8.0	10281	6 L	10552		C2502002	VL042052.D Seq :VL022625 TUNE : VL042049.D ICAL : VL042046.D CCAL : VL042050.D MB : VL042051.D
03/31/2025	B2502051	Q1665-14	-7.1	10275	6 L	10482		C2502003	VL042078.D Seq :VL030325 TUNE : VL042074.D ICAL : VL042046.D CCAL : VL042075.D MB : VL042076.D
03/31/2025	B2502051	Q1665-12	-13.7	10271	6 L	10222		C2502002	VL042052.D Seq :VL022625 TUNE : VL042049.D ICAL : VL042046.D CCAL : VL042050.D MB : VL042051.D
03/31/2025	B2502051	Q1665-04	-7.1	10268	6 L	10101		C2501003	VL041954.D Seq :VL012725 TUNE : VL041943.D ICAL : VL041894.D CCAL : VL041944.D MB : VL041946.D
03/31/2025	B2502051	Q1665-15	-6.3	10266	6 L	10696		C2503003	VL042145.D Seq :VL031425 TUNE : VL042140.D ICAL : VL042046.D CCAL : VL042141.D MB : VL042142.D
03/31/2025	B2502051	Q1665-06	-12.0	10198	6 L	10204		C2502002	VL042052.D Seq :VL022625 TUNE : VL042049.D ICAL : VL042046.D CCAL : VL042050.D MB : VL042051.D

Process Report

Order ID : Q1665

Date	Bottle ID	Lab Sample Number	InComing Vacuum	Canister ID	Can Size	Reg ID	Reg Size	Batch Code	Certification ID
03/31/2025	B2502051	Q1665-03	-8.4	10154	6 L	10476		C2502001	VL041974.D Seq :vl020325 TUNE : VL041957.D ICAL : VL041894.D CCAL : VL041958.D MB : VL041960.D
03/31/2025	B2502051	Q1665-08	-7.5	10053	6 L	10168		C2502001	VL041974.D Seq :vl020325 TUNE : VL041957.D ICAL : VL041894.D CCAL : VL041958.D MB : VL041960.D

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042041.D
 Acq On : 25 Feb 2025 08:56
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 03/06/2025
 Supervised By :Semsettin Yesilyurt 03/06/2025

Quant Time: Feb 25 10:10:59 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Feb 18 00:13:37 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	161135	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.978	114	422655	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.901	117	374249	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.393	95	285943	9.263	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.600%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	152005	5.480	ppbv	99
3) Chlorodifluoromethane	1.479	51	211185	6.887	ppbv	96
4) Chloromethane	1.538	50	55364	5.136	ppbv	98
5) Vinyl Chloride	1.586	62	54228	4.207	ppbv	98
6) Bromomethane	1.674	94	25846	5.173	ppbv	90
7) Chloroethane	1.709	64	21717	5.169	ppbv	96
8) Dichlorotetrafluoroethane	1.560	85	125595	5.147	ppbv	94
9) Propene	1.489	41	84038	6.357	ppbv	97
10) Heptane	5.007	43	267860	7.296	ppbv	97
11) Trichlorofluoromethane	1.884	101	152112	5.449	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.146	101	112374	5.162	ppbv	93
13) Ethanol	1.729	45	3178	2.131	ppbv	85
14) Bromoethene	1.787	108	42461	5.223	ppbv	99
15) Acetone	1.842	43	115085	4.327	ppbv	95
16) 1,3-Butadiene	1.615	39	57707	4.574	ppbv	100
17) tert-Butyl alcohol	2.049	59	153073	4.651	ppbv	100
18) 1,1-Dichloroethene	2.043	96	51534	4.954	ppbv	95
19) Isopropyl Alcohol	1.894	45	67445	4.522	ppbv	98
20) Methylene Chloride	2.069	84	41998	4.195	ppbv	94
21) Allyl Chloride	2.104	41	86723	4.504	ppbv	99
22) trans-1,2-Dichloroethene	2.350	96	57617	5.104	ppbv	96
23) Vinyl Acetate	2.473	43	83866	5.823	ppbv	98
24) 1,1-Dichloroethane	2.418	63	107817	4.704	ppbv	100
25) Ethyl Acetate	2.852	43	511505	7.007	ppbv	99
26) Hexane	2.839	57	206577	7.052	ppbv	98
27) Carbon Disulfide	2.159	76	137978	5.008	ppbv	97
28) Methyl tert-Butyl Ether	2.450	73	84091	4.696	ppbv	98
29) Chloroform	2.862	83	277411	6.538	ppbv	96
30) Cyclohexane	3.875	84	173351	7.295	ppbv	98
31) cis-1,2-Dichloroethene	2.732	61	199354	6.886	ppbv	98
32) 1,1,1-Trichloroethane	3.383	97	285335	6.064	ppbv	98
34) 2-Butanone	2.570	43	292449	8.043	ppbv	100
35) Carbon Tetrachloride	3.781	117	276677	6.741	ppbv	97
36) Benzene	3.674	78	409501	7.999	ppbv	98
37) 1,2-Dichloroethane	3.234	62	219829	7.543	ppbv	100
38) Trichloroethene	4.570	130	159933	7.689	ppbv	93
39) 1,2-Dichloropropane	4.337	63	151036	8.047	ppbv	99
40) 1,4-Dioxane	4.609	88	69399	7.105	ppbv #	76
41) Tetrahydrofuran	3.072	42	176830	8.509	ppbv	95
42) Bromodichloromethane	4.512	83	312250	7.908	ppbv	97
43) Methyl Methacrylate	4.907	69	165494	7.091	ppbv	97
44) 2,2,4-Trimethylpentane	4.661	57	705920	8.122	ppbv	97
45) t-1,3-Dichloropropene	6.444	75	166207	7.544	ppbv	91

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042041.D
 Acq On : 25 Feb 2025 08:56
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 03/06/2025
 Supervised By :Semsettin Yesilyurt 03/06/2025

Quant Time: Feb 25 10:10:59 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Feb 18 00:13:37 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.629	75	218294	7.809	ppbv	99
47) 1,1,2-Trichloroethane	6.610	97	152556	7.723	ppbv	97
48) Dibromochloromethane	7.412	129	251213	7.593	ppbv	99
49) Bromoform	9.503	173	222057	8.183	ppbv	98
50) 4-Methyl-2-Pentanone	5.787	43	428434	7.666	ppbv	100
51) 2-Hexanone	7.464	43	352331	7.709	ppbv #	96
52) Tetrachloroethene	8.244	164	138263	7.743	ppbv	94
53) Toluene	6.959	91	460041	7.541	ppbv	97
54) 1,2-Dibromoethane	7.678	107	211395	7.772	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.943	131	188811	7.049	ppbv	97
57) Chlorobenzene	8.940	112	340525	7.735	ppbv	97
58) Ethyl Benzene	9.374	91	633939	7.739	ppbv	98
59) m/p-Xylene	9.552	91	995259m	15.427	ppbv	
60) o-Xylene	9.982	91	493069	7.557	ppbv	97
61) Styrene	9.888	104	246161	7.912	ppbv	98
62) Isopropylbenzene	10.555	105	736884	7.522	ppbv	98
63) 1,1,2,2-Tetrachloroethane	9.979	83	317313	7.538	ppbv	100
64) n-propylbenzene	11.005	120	196044	8.145	ppbv	90
65) tert-Butylbenzene	11.545	119	648698	7.596	ppbv	97
66) Benzyl Chloride	11.630	91	57345	6.699	ppbv #	97
67) sec-Butylbenzene	11.782	105	928863	7.744	ppbv	99
69) p-Isopropyltoluene	11.940	119	756264	7.702	ppbv	98
70) n-Butylbenzene	12.296	91	774646	8.009	ppbv	99
71) 2-Chlorotoluene	10.918	91	566619	7.681	ppbv	97
72) 4-Ethyltoluene	11.141	105	606538	7.934	ppbv	98
73) 1,3,5-Trimethylbenzene	11.219	105	516875	7.838	ppbv	100
74) 1,2,4-Trimethylbenzene	11.552	105	550412	7.777	ppbv	98
75) 1,3-Dichlorobenzene	11.623	146	325160	8.535	ppbv	97
76) 1,4-Dichlorobenzene	11.685	146	322404	8.678	ppbv	97
77) 1,2-Dichlorobenzene	11.963	146	312942	8.593	ppbv	96
78) Hexachloro-1,3-Butadiene	13.895	225	238217	7.907	ppbv	100
79) Naphthalene	13.526	128	513192	7.372	ppbv	99
80) Naphthalene,2-methyl-	14.471	142	210405	6.676	ppbv	95
81) 1,2,4-Trichlorobenzene	13.455	180	241031	8.517	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Quant Time: Feb 25 10:10:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Tue Feb 18 00:13:37 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042042.D
 Acq On : 25 Feb 2025 09:29
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
 Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 10:12:18 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Feb 18 00:13:37 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	160999	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.981	114	427051	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.901	117	377910	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.393	95	289049	9.273	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.505	85	35122	1.267	ppbv	95
3) Chlorodifluoromethane	1.483	51	49423	1.613	ppbv	94
4) Chloromethane	1.538	50	12833	1.192	ppbv #	87
5) Vinyl Chloride	1.586	62	11611	0.902	ppbv	97
6) Bromomethane	1.674	94	5996	1.201	ppbv	97
7) Chloroethane	1.712	64	5014	1.195	ppbv	90
8) Dichlorotetrafluoroethane	1.560	85	29028	1.191	ppbv	99
9) Propene	1.489	41	20484	1.551	ppbv	95
10) Heptane	5.004	43	63098	1.720	ppbv	94
11) Trichlorofluoromethane	1.884	101	36041	1.292	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.149	101	27846	1.280	ppbv	90
13) Ethanol	1.729	45	1291	0.866	ppbv	91
14) Bromoethene	1.790	108	9642	1.187	ppbv	93
15) Acetone	1.842	43	30058	1.131	ppbv	100
16) 1,3-Butadiene	1.615	39	13471	1.069	ppbv	100
17) tert-Butyl alcohol	2.052	59	38684	1.176	ppbv	96
18) 1,1-Dichloroethene	2.043	96	12417	1.195	ppbv	93
19) Isopropyl Alcohol	1.894	45	17927	1.203	ppbv	96
20) Methylene Chloride	2.072	84	10587	1.058	ppbv #	80
21) Allyl Chloride	2.107	41	20564	1.069	ppbv	96
22) trans-1,2-Dichloroethene	2.350	96	14475	1.283	ppbv	81
23) Vinyl Acetate	2.473	43	22394	1.556	ppbv #	94
24) 1,1-Dichloroethane	2.421	63	27804	1.214	ppbv	99
25) Ethyl Acetate	2.852	43	121593	1.667	ppbv	100
26) Hexane	2.839	57	48912	1.671	ppbv	98
27) Carbon Disulfide	2.159	76	33222	1.207	ppbv #	57
28) Methyl tert-Butyl Ether	2.457	73	26968	1.507	ppbv	100
29) Chloroform	2.865	83	63463	1.497	ppbv	92
30) Cyclohexane	3.878	84	40420	1.702	ppbv	98
31) cis-1,2-Dichloroethene	2.732	61	46690	1.614	ppbv	96
32) 1,1,1-Trichloroethane	3.386	97	62586	1.331	ppbv	97
34) 2-Butanone	2.573	43	69370	1.888	ppbv	96
35) Carbon Tetrachloride	3.784	117	61895	1.493	ppbv	97
36) Benzene	3.677	78	94277	1.822	ppbv	98
37) 1,2-Dichloroethane	3.237	62	51050	1.734	ppbv	99
38) Trichloroethene	4.574	130	38184	1.817	ppbv	99
39) 1,2-Dichloropropane	4.341	63	35396	1.866	ppbv	95
40) 1,4-Dioxane	4.622	88	15442	1.565	ppbv #	28
41) Tetrahydrofuran	3.075	42	40365	1.922	ppbv	96
42) Bromodichloromethane	4.512	83	69154	1.733	ppbv	94
43) Methyl Methacrylate	4.910	69	36393	1.543	ppbv	94
44) 2,2,4-Trimethylpentane	4.664	57	166945	1.901	ppbv	97
45) t-1,3-Dichloropropene	6.448	75	35707	1.604	ppbv	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042042.D
 Acq On : 25 Feb 2025 09:29
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
 Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 10:12:18 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 18 00:13:37 2025


Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.632	75	46835	1.658	ppbv	93
47) 1,1,2-Trichloroethane	6.613	97	36368	1.822	ppbv	93
48) Dibromochloromethane	7.412	129	56931	1.703	ppbv	99
49) Bromoform	9.503	173	50259	1.833	ppbv	99
50) 4-Methyl-2-Pentanone	5.797	43	83246	1.474	ppbv	99
51) 2-Hexanone	7.470	43	51926	1.124	ppbv #	94
52) Tetrachloroethene	8.244	164	32874	1.822	ppbv	95
53) Toluene	6.962	91	109646	1.779	ppbv	98
54) 1,2-Dibromoethane	7.678	107	50010	1.820	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.946	131	44947	1.662	ppbv	98
57) Chlorobenzene	8.943	112	84749	1.906	ppbv	95
58) Ethyl Benzene	9.374	91	149185	1.804	ppbv	99
59) m/p-Xylene	9.568	91	238800m	3.666	ppbv	
60) o-Xylene	9.982	91	119170	1.809	ppbv	97
61) Styrene	9.888	104	55798	1.776	ppbv	96
62) Isopropylbenzene	10.555	105	181311	1.833	ppbv	97
63) 1,1,2,2-Tetrachloroethane	9.982	83	77911	1.833	ppbv	96
64) n-propylbenzene	11.005	120	47764	1.965	ppbv	92
65) tert-Butylbenzene	11.545	119	167973	1.948	ppbv	96
66) Benzyl Chloride	11.630	91	12304	1.423	ppbv #	98
67) sec-Butylbenzene	11.782	105	240504	1.986	ppbv	99
69) p-Isopropyltoluene	11.940	119	194660	1.963	ppbv	99
70) n-Butylbenzene	12.296	91	194416	1.991	ppbv	99
71) 2-Chlorotoluene	10.917	91	140068	1.880	ppbv	98
72) 4-Ethyltoluene	11.141	105	150038	1.944	ppbv	98
73) 1,3,5-Trimethylbenzene	11.218	105	128202	1.925	ppbv	99
74) 1,2,4-Trimethylbenzene	11.549	105	150380	2.104	ppbv #	88
75) 1,3-Dichlorobenzene	11.623	146	81884	2.129	ppbv	98
76) 1,4-Dichlorobenzene	11.688	146	82560	2.201	ppbv	97
77) 1,2-Dichlorobenzene	11.963	146	80367	2.186	ppbv	96
78) Hexachloro-1,3-Butadiene	13.895	225	63815	2.098	ppbv	99
79) Naphthalene	13.526	128	102197	1.454	ppbv	98
80) Naphthalene,2-methyl-	14.475	142	18104	0.569	ppbv	94
81) 1,2,4-Trichlorobenzene	13.455	180	57798	2.023	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Quant Time: Feb 25 10:12:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Tue Feb 18 00:13:37 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042043.D
 Acq On : 25 Feb 2025 10:00
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 02/26/2025
 Supervised By : Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 10:27:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 25 10:15:00 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	161096	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.978	114	423363	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.898	117	369917	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.390	95	287450	9.420	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	17457	0.629	ppbv	95
3) Chlorodifluoromethane	1.479	51	25229	0.823	ppbv	98
4) Chloromethane	1.538	50	6417	0.595	ppbv #	75
5) Vinyl Chloride	1.586	62	5567	0.432	ppbv	96
6) Bromomethane	1.674	94	2966	0.594	ppbv	100
7) Chloroethane	1.709	64	2542	0.605	ppbv	96
8) Dichlorotetrafluoroethane	1.560	85	14655	0.601	ppbv	90
9) Propene	1.486	41	10156	0.768	ppbv	96
10) Heptane	4.998	43	31354m	0.854	ppbv	
11) Trichlorofluoromethane	1.884	101	18079	0.648	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.149	101	14527	0.667	ppbv	92
13) Ethanol	1.725	45	915m	0.614	ppbv	
14) Bromoethene	1.787	108	4905	0.604	ppbv #	95
15) Acetone	1.845	43	15468	0.582	ppbv	98
16) 1,3-Butadiene	1.615	39	7010	0.556	ppbv	99
17) tert-Butyl alcohol	2.052	59	19092	0.580	ppbv	97
18) 1,1-Dichloroethene	2.043	96	7042	0.677	ppbv	94
19) Isopropyl Alcohol	1.894	45	9191	0.616	ppbv #	32
20) Methylene Chloride	2.068	84	6026	0.602	ppbv	95
21) Allyl Chloride	2.104	41	11550	0.600	ppbv	95
22) trans-1,2-Dichloroethene	2.347	96	8145m	0.722	ppbv	
23) Vinyl Acetate	2.473	43	9893	0.687	ppbv #	91
24) 1,1-Dichloroethane	2.418	63	13783	0.601	ppbv	98
25) Ethyl Acetate	2.852	43	59183	0.811	ppbv	99
26) Hexane	2.836	57	23549	0.804	ppbv	96
27) Carbon Disulfide	2.159	76	15659	0.568	ppbv #	46
28) Methyl tert-Butyl Ether	2.454	73	13593	0.759	ppbv	95
29) Chloroform	2.861	83	32925	0.776	ppbv	94
30) Cyclohexane	3.874	84	19953	0.840	ppbv	97
31) cis-1,2-Dichloroethene	2.732	61	22565	0.780	ppbv	97
32) 1,1,1-Trichloroethane	3.386	97	30870	0.656	ppbv	96
34) 2-Butanone	2.570	43	33820	0.929	ppbv	99
35) Carbon Tetrachloride	3.777	117	30618	0.745	ppbv	90
36) Benzene	3.674	78	45892	0.895	ppbv	96
37) 1,2-Dichloroethane	3.234	62	25585	0.876	ppbv	99
38) Trichloroethene	4.570	130	18921	0.908	ppbv	91
39) 1,2-Dichloropropane	4.334	63	17557	0.934	ppbv	94
40) 1,4-Dioxane	4.619	88	8319m	0.850	ppbv	
41) Tetrahydrofuran	3.075	42	20823	1.000	ppbv	96
42) Bromodichloromethane	4.515	83	33481	0.847	ppbv	97
43) Methyl Methacrylate	4.907	69	17679	0.756	ppbv	99
44) 2,2,4-Trimethylpentane	4.661	57	81921	0.941	ppbv	98
45) t-1,3-Dichloropropene	6.444	75	16578m	0.751	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042043.D
 Acq On : 25 Feb 2025 10:00
 Operator : SY/MD
 Sample : VSTDIC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
 Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 10:27:35 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 25 10:15:00 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.632	75	22127m	0.790	ppbv	
47) 1,1,2-Trichloroethane	6.616	97	18160m	0.918	ppbv	
48) Dibromochloromethane	7.412	129	28239	0.852	ppbv	97
49) Bromoform	9.500	173	23310	0.858	ppbv	94
50) 4-Methyl-2-Pentanone	5.791	43	41249m	0.737	ppbv	
51) 2-Hexanone	7.470	43	24117	0.527	ppbv #	97
52) Tetrachloroethene	8.241	164	16217	0.907	ppbv	89
53) Toluene	6.959	91	51231	0.838	ppbv	99
54) 1,2-Dibromoethane	7.674	107	23427	0.860	ppbv	94
56) 1,1,1,2-Tetrachloroethane	8.943	131	22060	0.833	ppbv	97
57) Chlorobenzene	8.937	112	41173	0.946	ppbv	96
58) Ethyl Benzene	9.370	91	70507	0.871	ppbv	97
59) m/p-Xylene	9.568	91	114691m	1.799	ppbv	
60) o-Xylene	9.979	91	57322	0.889	ppbv	99
61) Styrene	9.885	104	26596	0.865	ppbv	98
62) Isopropylbenzene	10.555	105	86533	0.894	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.979	83	38520	0.926	ppbv	98
64) n-propylbenzene	11.002	120	22456	0.944	ppbv	94
65) tert-Butylbenzene	11.545	119	83284	0.987	ppbv	99
66) Benzyl Chloride	11.626	91	5880m	0.695	ppbv	
67) sec-Butylbenzene	11.782	105	119435	1.007	ppbv	99
69) p-Isopropyltoluene	11.940	119	93710	0.966	ppbv	99
70) n-Butylbenzene	12.293	91	94969	0.993	ppbv	98
71) 2-Chlorotoluene	10.914	91	67655	0.928	ppbv	98
72) 4-Ethyltoluene	11.137	105	71906	0.952	ppbv	99
73) 1,3,5-Trimethylbenzene	11.215	105	64087	0.983	ppbv	96
74) 1,2,4-Trimethylbenzene	11.549	105	73323	1.048	ppbv	100
75) 1,3-Dichlorobenzene	11.620	146	41063	1.091	ppbv	99
76) 1,4-Dichlorobenzene	11.685	146	39235	1.068	ppbv	99
77) 1,2-Dichlorobenzene	11.960	146	40574	1.127	ppbv	98
78) Hexachloro-1,3-Butadiene	13.895	225	30764	1.033	ppbv	99
79) Naphthalene	13.526	128	45743	0.665	ppbv	98
80) Naphthalene,2-methyl-	14.471	142	5244	0.168	ppbv	94
81) 1,2,4-Trichlorobenzene	13.452	180	25404	0.908	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042044.D
 Acq On : 25 Feb 2025 10:31
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Quant Time: Feb 25 10:56:41 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 25 10:31:48 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 02/26/2025
 Supervised By : Mahesh Dadoda 02/26/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.797	49	159644	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.975	114	421340	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.898	117	365954	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.390	95	281042	9.310	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	8852	0.322	ppbv	98
3) Chlorodifluoromethane	1.479	51	11921	0.392	ppbv	94
4) Chloromethane	1.537	50	3497	0.327	ppbv #	88
5) Vinyl Chloride	1.583	62	3005	0.235	ppbv	98
6) Bromomethane	1.673	94	1359	0.275	ppbv	92
7) Chloroethane	1.709	64	1205	0.290	ppbv	89
8) Dichlorotetrafluoroethane	1.557	85	7123	0.295	ppbv	99
9) Propene	1.486	41	4904	0.374	ppbv	94
10) Heptane	5.004	43	15731	0.432	ppbv	98
11) Trichlorofluoromethane	1.884	101	8429	0.305	ppbv	92
12) 1,1,2-Trichlorotrifluo...	2.146	101	7336	0.340	ppbv	98
13) Ethanol	1.725	45	426	0.288	ppbv	75
14) Bromoethene	1.787	108	2598	0.323	ppbv #	89
15) Acetone	1.842	43	8655	0.328	ppbv	96
16) 1,3-Butadiene	1.615	39	3803	0.304	ppbv	97
17) tert-Butyl alcohol	2.049	59	9447	0.290	ppbv #	90
18) 1,1-Dichloroethene	2.039	96	3173	0.308	ppbv	94
19) Isopropyl Alcohol	1.894	45	4481	0.303	ppbv #	32
20) Methylene Chloride	2.068	84	3423	0.345	ppbv	97
21) Allyl Chloride	2.104	41	5640	0.296	ppbv	100
22) trans-1,2-Dichloroethene	2.347	96	3838m	0.343	ppbv	
23) Vinyl Acetate	2.473	43	5499	0.385	ppbv #	86
24) 1,1-Dichloroethane	2.418	63	7396	0.326	ppbv #	88
25) Ethyl Acetate	2.848	43	30593	0.423	ppbv	98
26) Hexane	2.832	57	12739	0.439	ppbv	98
27) Carbon Disulfide	2.159	76	8316	0.305	ppbv #	15
28) Methyl tert-Butyl Ether	2.450	73	5185	0.292	ppbv #	80
29) Chloroform	2.858	83	17278	0.411	ppbv	93
30) Cyclohexane	3.874	84	9209	0.391	ppbv	88
31) cis-1,2-Dichloroethene	2.729	61	11344	0.395	ppbv	97
32) 1,1,1-Trichloroethane	3.382	97	15027	0.322	ppbv	93
34) 2-Butanone	2.570	43	17555	0.484	ppbv	99
35) Carbon Tetrachloride	3.780	117	14223	0.348	ppbv	90
36) Benzene	3.670	78	22728	0.445	ppbv	95
37) 1,2-Dichloroethane	3.230	62	13141	0.452	ppbv	100
38) Trichloroethene	4.570	130	9115	0.440	ppbv #	87
39) 1,2-Dichloropropane	4.331	63	8877m	0.474	ppbv	
40) 1,4-Dioxane	4.622	88	4046m	0.416	ppbv	
41) Tetrahydrofuran	3.078	42	9880	0.477	ppbv	98
42) Bromodichloromethane	4.505	83	16973	0.431	ppbv #	96
43) Methyl Methacrylate	4.904	69	8428	0.362	ppbv	93
44) 2,2,4-Trimethylpentane	4.661	57	40623	0.469	ppbv	98
45) t-1,3-Dichloropropene	6.441	75	8720m	0.397	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042044.D
 Acq On : 25 Feb 2025 10:31
 Operator : SY/MD
 Sample : VSTDICC0.5
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.5

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
 Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 10:56:41 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 25 10:31:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.629	75	10834m	0.389	ppbv	
47) 1,1,2-Trichloroethane	6.609	97	8982m	0.456	ppbv	
48) Dibromochloromethane	7.406	129	13014m	0.395	ppbv	
49) Bromoform	9.496	173	11640	0.430	ppbv	98
50) 4-Methyl-2-Pentanone	5.794	43	20992m	0.377	ppbv	
51) 2-Hexanone	7.470	43	12149m	0.267	ppbv	
52) Tetrachloroethene	8.241	164	8012	0.450	ppbv	95
53) Toluene	6.956	91	24224	0.398	ppbv	96
54) 1,2-Dibromoethane	7.674	107	11718m	0.432	ppbv	
56) 1,1,1,2-Tetrachloroethane	8.940	131	11100	0.424	ppbv #	61
57) Chlorobenzene	8.936	112	20167	0.468	ppbv	97
58) Ethyl Benzene	9.370	91	34506	0.431	ppbv	90
59) m/p-Xylene	9.568	91	54449m	0.863	ppbv	
60) o-Xylene	9.979	91	28757	0.451	ppbv	95
61) Styrene	9.885	104	12051	0.396	ppbv	99
62) Isopropylbenzene	10.552	105	42074	0.439	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.975	83	18774	0.456	ppbv	96
64) n-propylbenzene	11.001	120	10900	0.463	ppbv	100
65) tert-Butylbenzene	11.542	119	38066	0.456	ppbv	98
66) Benzyl Chloride	11.626	91	2589	0.309	ppbv #	96
67) sec-Butylbenzene	11.778	105	54394	0.464	ppbv	97
69) p-Isopropyltoluene	11.937	119	40978	0.427	ppbv	95
70) n-Butylbenzene	12.293	91	42673	0.451	ppbv	99
71) 2-Chlorotoluene	10.914	91	32161	0.446	ppbv	99
72) 4-Ethyltoluene	11.137	105	32477	0.434	ppbv #	95
73) 1,3,5-Trimethylbenzene	11.215	105	29130	0.452	ppbv	91
74) 1,2,4-Trimethylbenzene	11.548	105	33950	0.491	ppbv	97
75) 1,3-Dichlorobenzene	11.616	146	20339	0.546	ppbv	96
76) 1,4-Dichlorobenzene	11.684	146	20492	0.564	ppbv	93
77) 1,2-Dichlorobenzene	11.959	146	19373	0.544	ppbv	96
78) Hexachloro-1,3-Butadiene	13.892	225	15367	0.522	ppbv	99
79) Naphthalene	13.523	128	17897	0.263	ppbv #	95
80) Naphthalene,2-methyl-	14.478	142	1646	0.053	ppbv #	81
81) 1,2,4-Trichlorobenzene	13.452	180	10771	0.389	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042045.D
 Acq On : 25 Feb 2025 11:03
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
 Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 12:33:08 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Tue Feb 25 10:31:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.803	49	161406	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.981	114	414875	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.904	117	363831	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.393	95	276484	9.213	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.100%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.586	62	543	0.042	ppbv #	88
32) 1,1,1-Trichloroethane	3.386	97	3026	0.064	ppbv #	86
35) Carbon Tetrachloride	3.787	117	2923m	0.073	ppbv	
38) Trichloroethene	4.577	130	1915m	0.094	ppbv	
52) Tetrachloroethene	8.254	164	1552m	0.089	ppbv	
54) 1,2-Dibromoethane	7.678	107	2335m	0.087	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.982	83	3549m	0.087	ppbv	
79) Naphthalene	13.529	128	3058	0.045	ppbv #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
Data File : VL042045.D
Acq On : 25 Feb 2025 11:03
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 12:33:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Feb 25 10:31:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042046.D
 Acq On : 25 Feb 2025 11:34
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
 Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 12:34:19 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 25 10:31:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.797	49	160156	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.975	114	410900	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.898	117	353154	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.390	95	272906	9.368	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.700%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.586	62	215	0.017	ppbv	89
32) 1,1,1-Trichloroethane	3.379	97	923m	0.020	ppbv	
35) Carbon Tetrachloride	3.784	117	818m	0.021	ppbv	
38) Trichloroethene	4.577	130	551m	0.027	ppbv	
52) Tetrachloroethene	8.237	164	397m	0.023	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.979	83	1058m	0.027	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
Data File : VL042046.D
Acq On : 25 Feb 2025 11:34
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 02/26/2025
Supervised By : Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 12:34:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Feb 25 10:31:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042047.D
 Acq On : 25 Feb 2025 12:07
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
 Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 12:35:10 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 25 10:31:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	159015	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.978	114	401919	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.904	117	351637	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.393 95 273846 9.441 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 94.400%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	231978	8.474	ppbv	99
3) Chlorodifluoromethane	1.479	51	324821	10.733	ppbv	100
4) Chloromethane	1.537	50	82097	7.718	ppbv	98
5) Vinyl Chloride	1.586	62	81863	6.436	ppbv	99
6) Bromomethane	1.673	94	39366	7.983	ppbv	97
7) Chloroethane	1.709	64	33329	8.039	ppbv	92
8) Dichlorotetrafluoroethane	1.560	85	188716	7.837	ppbv	98
9) Propene	1.489	41	133086	10.201	ppbv	97
10) Heptane	5.004	43	400418	11.052	ppbv	99
11) Trichlorofluoromethane	1.884	101	230938	8.383	ppbv	95
12) 1,1,2-Trichlorotrifluo...	2.149	101	184306	8.579	ppbv	99
13) Ethanol	1.725	45	4281	2.909	ppbv	100
14) Bromoethene	1.787	108	65496	8.164	ppbv	99
15) Acetone	1.842	43	172538	6.574	ppbv	98
16) 1,3-Butadiene	1.615	39	87490	7.027	ppbv	100
17) tert-Butyl alcohol	2.049	59	239569	7.377	ppbv	98
18) 1,1-Dichloroethene	2.042	96	80535	7.845	ppbv	98
19) Isopropyl Alcohol	1.894	45	102365	6.955	ppbv	99
20) Methylene Chloride	2.068	84	66282	6.709	ppbv	98
21) Allyl Chloride	2.104	41	140726	7.406	ppbv	99
22) trans-1,2-Dichloroethene	2.350	96	94321	8.466	ppbv	95
23) Vinyl Acetate	2.473	43	139727	9.831	ppbv	99
24) 1,1-Dichloroethane	2.418	63	180882	7.997	ppbv	99
25) Ethyl Acetate	2.848	43	767353	10.652	ppbv	99
26) Hexane	2.839	57	306864	10.615	ppbv	99
27) Carbon Disulfide	2.159	76	224609	8.261	ppbv	98
28) Methyl tert-Butyl Ether	2.453	73	168058	9.510	ppbv	98
29) Chloroform	2.861	83	409240	9.774	ppbv	100
30) Cyclohexane	3.874	84	256687	10.946	ppbv	98
31) cis-1,2-Dichloroethene	2.732	61	299324	10.476	ppbv	96
32) 1,1,1-Trichloroethane	3.386	97	417615	8.994	ppbv	98
34) 2-Butanone	2.570	43	457602	13.234	ppbv	98
35) Carbon Tetrachloride	3.780	117	410730	10.524	ppbv	96
36) Benzene	3.674	78	602107	12.367	ppbv	97
37) 1,2-Dichloroethane	3.233	62	335275	12.097	ppbv	99
38) Trichloroethene	4.573	130	228298	11.542	ppbv	98
39) 1,2-Dichloropropane	4.337	63	228418	12.797	ppbv	98
40) 1,4-Dioxane	4.609	88	104519	11.253	ppbv #	96
41) Tetrahydrofuran	3.068	42	271202	13.724	ppbv	100
42) Bromodichloromethane	4.512	83	461272	12.285	ppbv	96
43) Methyl Methacrylate	4.907	69	246468	11.105	ppbv	97
44) 2,2,4-Trimethylpentane	4.667	57	1054880	12.763	ppbv	99
45) t-1,3-Dichloropropene	6.444	75	249612	11.914	ppbv	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042047.D
 Acq On : 25 Feb 2025 12:07
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
 Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 12:35:10 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 25 10:31:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.635	75	317266	11.935	ppbv	98
47) 1,1,2-Trichloroethane	6.609	97	226021	12.032	ppbv	95
48) Dibromochloromethane	7.412	129	369666	11.750	ppbv	99
49) Bromoform	9.503	173	317322	12.297	ppbv	99
50) 4-Methyl-2-Pentanone	5.787	43	650307	12.237	ppbv	99
51) 2-Hexanone	7.461	43	530165	12.198	ppbv #	100
52) Tetrachloroethene	8.247	164	203233	11.968	ppbv	94
53) Toluene	6.959	91	673822	11.615	ppbv	99
54) 1,2-Dibromoethane	7.677	107	309982	11.985	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.946	131	276326	10.979	ppbv	99
57) Chlorobenzene	8.943	112	494606	11.957	ppbv	98
58) Ethyl Benzene	9.373	91	922265	11.984	ppbv	100
59) m/p-Xylene	9.571	91	1456036m	24.020	ppbv	
60) o-Xylene	9.982	91	712735	11.626	ppbv	99
61) Styrene	9.888	104	363855	12.446	ppbv	99
62) Isopropylbenzene	10.558	105	1068777	11.612	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.982	83	459839	11.626	ppbv	100
64) n-propylbenzene	11.005	120	282364	12.486	ppbv	99
65) tert-Butylbenzene	11.548	119	924810	11.526	ppbv	99
66) Benzyl Chloride	11.633	91	78975	9.819	ppbv	98
67) sec-Butylbenzene	11.785	105	1315730	11.674	ppbv	99
69) p-Isopropyltoluene	11.940	119	1084253	11.752	ppbv	99
70) n-Butylbenzene	12.296	91	1118373	12.306	ppbv	99
71) 2-Chlorotoluene	10.917	91	818204	11.804	ppbv	100
72) 4-Ethyltoluene	11.141	105	880883	12.264	ppbv	99
73) 1,3,5-Trimethylbenzene	11.218	105	738619	11.921	ppbv	98
74) 1,2,4-Trimethylbenzene	11.552	105	784701	11.801	ppbv	91
75) 1,3-Dichlorobenzene	11.623	146	465436	13.003	ppbv	98
76) 1,4-Dichlorobenzene	11.688	146	467249	13.386	ppbv	99
77) 1,2-Dichlorobenzene	11.963	146	449222	13.129	ppbv	98
78) Hexachloro-1,3-Butadiene	13.895	225	351085	12.403	ppbv	100
79) Naphthalene	13.526	128	764990	11.696	ppbv	99
80) Naphthalene,2-methyl-	14.471	142	324431	10.956	ppbv	99
81) 1,2,4-Trichlorobenzene	13.455	180	354990	13.351	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022625\
 Data File : VL042050.D
 Acq On : 26 Feb 2025 08:38
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/27/2025
 Supervised By :Mahesh Dadoda 02/27/2025

Quant Time: Feb 27 00:24:17 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Feb 26 01:08:36 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.797	49	157492	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.975	114	387144	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.898	117	336136	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.390 95 264623 10.233 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.300%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	159125	9.716	ppbv	96
3) Chlorodifluoromethane	1.479	51	217238	9.491	ppbv	99
4) Chloromethane	1.538	50	54275	8.962	ppbv	97
5) Vinyl Chloride	1.583	62	53871	9.404	ppbv	99
6) Bromomethane	1.670	94	25674	9.412	ppbv	100
7) Chloroethane	1.709	64	22033	9.466	ppbv	89
8) Dichlorotetrafluoroethane	1.557	85	129383	9.610	ppbv	99
9) Propene	1.486	41	87439	9.377	ppbv	97
10) Heptane	5.001	43	268030	9.231	ppbv	99
11) Trichlorofluoromethane	1.884	101	155832	9.496	ppbv	98
12) 1,1,2-Trichlorotrifluo...	2.146	101	114341	8.735	ppbv	99
13) Ethanol	1.725	45	2918	4.929	ppbv	98
14) Bromoethene	1.787	108	44414	9.608	ppbv	95
15) Acetone	1.842	43	116372	8.367	ppbv	98
16) 1,3-Butadiene	1.612	39	59133	9.136	ppbv	99
17) tert-Butyl alcohol	2.046	59	153167	8.802	ppbv	96
18) 1,1-Dichloroethene	2.039	96	51101	8.640	ppbv	93
19) Isopropyl Alcohol	1.890	45	69014	8.638	ppbv	98
20) Methylene Chloride	2.068	84	42907	8.156	ppbv	98
21) Allyl Chloride	2.101	41	86423	8.600	ppbv	99
22) trans-1,2-Dichloroethene	2.347	96	61551	8.927	ppbv	98
23) Vinyl Acetate	2.470	43	84867	8.678	ppbv	98
24) 1,1-Dichloroethane	2.418	63	118286	9.220	ppbv	99
25) Ethyl Acetate	2.848	43	521677	9.371	ppbv	99
26) Hexane	2.835	57	207168	9.205	ppbv	98
27) Carbon Disulfide	2.156	76	138544	9.081	ppbv	95
28) Methyl tert-Butyl Ether	2.447	73	88652	7.912	ppbv	97
29) Chloroform	2.858	83	273992	9.045	ppbv	98
30) Cyclohexane	3.871	84	167364	9.162	ppbv	94
31) cis-1,2-Dichloroethene	2.732	61	197466	9.268	ppbv	97
32) 1,1,1-Trichloroethane	3.379	97	275336	9.370	ppbv	97
34) 2-Butanone	2.567	43	298594	9.900	ppbv	97
35) Carbon Tetrachloride	3.777	117	272097	10.192	ppbv	97
36) Benzene	3.670	78	394494	9.734	ppbv	97
37) 1,2-Dichloroethane	3.230	62	224585	9.995	ppbv	99
38) Trichloroethene	4.567	130	149021	9.005	ppbv	97
39) 1,2-Dichloropropane	4.331	63	151626	9.857	ppbv	95
40) 1,4-Dioxane	4.606	88	68639	9.775	ppbv #	98
41) Tetrahydrofuran	3.065	42	178871	10.037	ppbv	97
42) Bromodichloromethane	4.509	83	306517	10.124	ppbv	98
43) Methyl Methacrylate	4.900	69	156460	9.885	ppbv	96
44) 2,2,4-Trimethylpentane	4.658	57	703370	9.834	ppbv	99
45) t-1,3-Dichloropropene	6.438	75	155197	9.870	ppbv	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022625\
 Data File : VL042050.D
 Acq On : 26 Feb 2025 08:38
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDCCC010

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/27/2025
 Supervised By :Mahesh Dadoda 02/27/2025

Quant Time: Feb 27 00:24:17 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Wed Feb 26 01:08:36 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.622	75	205249	10.087	ppbv	99
47) 1,1,2-Trichloroethane	6.603	97	148058	9.481	ppbv	96
48) Dibromochloromethane	7.409	129	240304	9.825	ppbv	100
49) Bromoform	9.496	173	205649	9.682	ppbv	98
50) 4-Methyl-2-Pentanone	5.778	43	423271	10.851	ppbv	97
51) 2-Hexanone	7.454	43	345562	12.872	ppbv #	100
52) Tetrachloroethene	8.237	164	131164	9.454	ppbv	97
53) Toluene	6.952	91	439634	9.706	ppbv	98
54) 1,2-Dibromoethane	7.671	107	200726	9.507	ppbv	97
56) 1,1,1,2-Tetrachloroethane	8.940	131	178871	9.415	ppbv	100
57) Chlorobenzene	8.937	112	326805	9.377	ppbv	99
58) Ethyl Benzene	9.367	91	598242	9.664	ppbv	99
59) m/p-Xylene	9.564	91	957021m	19.427	ppbv	
60) o-Xylene	9.979	91	475848	9.608	ppbv	100
61) Styrene	9.885	104	231133	9.927	ppbv	99
62) Isopropylbenzene	10.552	105	718826	9.692	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.975	83	312195	9.572	ppbv	98
64) n-propylbenzene	11.001	120	185068	9.513	ppbv	95
65) tert-Butylbenzene	11.542	119	618290	9.160	ppbv	98
66) Benzyl Chloride	11.629	91	49850	9.678	ppbv	98
67) sec-Butylbenzene	11.778	105	890927	9.229	ppbv	100
69) p-Isopropyltoluene	11.937	119	724678	9.435	ppbv	100
70) n-Butylbenzene	12.293	91	756366	9.648	ppbv	99
71) 2-Chlorotoluene	10.914	91	545538	9.541	ppbv	99
72) 4-Ethyltoluene	11.137	105	574661	9.504	ppbv	99
73) 1,3,5-Trimethylbenzene	11.215	105	484912	9.245	ppbv	98
74) 1,2,4-Trimethylbenzene	11.548	105	527932	8.938	ppbv	96
75) 1,3-Dichlorobenzene	11.620	146	309408	9.102	ppbv	99
76) 1,4-Dichlorobenzene	11.684	146	309423	9.168	ppbv	99
77) 1,2-Dichlorobenzene	11.960	146	302720	9.177	ppbv	99
78) Hexachloro-1,3-Butadiene	13.892	225	229379	8.937	ppbv	100
79) Naphthalene	13.523	128	493813	12.193	ppbv	99
80) Naphthalene,2-methyl-	14.468	142	190978	17.232	ppbv	99
81) 1,2,4-Trichlorobenzene	13.452	180	226561	10.038	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022625\
 Data File : VL042052.D
 Acq On : 26 Feb 2025 10:04
 Operator : SY/MD
 Sample : QC021725 CAN 10271
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC021725 CAN 10271

Quant Time: Feb 27 00:25:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Feb 26 01:08:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.794	49	152414	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.969	114	364923	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.891	117	316584	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.384	95	246219	10.109	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.100%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022625\
Data File : VL042052.D
Acq On : 26 Feb 2025 10:04
Operator : SY/MD
Sample : QC021725 CAN 10271
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC021725 CAN 10271

Quant Time: Feb 27 00:25:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Feb 26 01:08:36 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022625\
 Data File : VL042054.D
 Acq On : 26 Feb 2025 11:07
 Operator : SY/MD
 Sample : VL0226ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0226ABS01

Quant Time: Feb 27 00:26:21 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Feb 26 01:08:36 2025

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 02/27/2025
 Supervised By : Mahesh Dadoda 02/27/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.797	49	153181	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.972	114	369057	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.895	117	318885	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.387 95 257152 10.482 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 104.800%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	154942	9.727	ppbv	100
3) Chlorodifluoromethane	1.480	51	208345	9.358	ppbv	100
4) Chloromethane	1.538	50	53585	9.097	ppbv	99
5) Vinyl Chloride	1.583	62	54944	9.861	ppbv	96
6) Bromomethane	1.674	94	25364	9.560	ppbv	96
7) Chloroethane	1.709	64	21007	9.279	ppbv	88
8) Dichlorotetrafluoroethane	1.557	85	125091	9.553	ppbv	98
9) Propene	1.486	41	83500	9.206	ppbv	99
10) Heptane	5.001	43	256046	9.067	ppbv	99
11) Trichlorofluoromethane	1.884	101	154527	9.682	ppbv	97
12) 1,1,2-Trichlorotrifluo...	2.146	101	124837	9.805	ppbv	96
13) Ethanol	1.726	45	3000	5.211	ppbv	92
14) Bromoethene	1.787	108	43051	9.576	ppbv	99
15) Acetone	1.842	43	116404	8.605	ppbv	100
16) 1,3-Butadiene	1.612	39	57319	9.105	ppbv	99
17) tert-Butyl alcohol	2.049	59	156939	9.273	ppbv	99
18) 1,1-Dichloroethene	2.039	96	53389	9.281	ppbv	86
19) Isopropyl Alcohol	1.891	45	68899	8.867	ppbv	100
20) Methylene Chloride	2.069	84	46174	9.024	ppbv	93
21) Allyl Chloride	2.101	41	92702	9.484	ppbv	99
22) trans-1,2-Dichloroethene	2.347	96	61477	9.167	ppbv	99
23) Vinyl Acetate	2.470	43	80476	8.461	ppbv	98
24) 1,1-Dichloroethane	2.415	63	119185	9.551	ppbv	99
25) Ethyl Acetate	2.845	43	509260	9.405	ppbv	100
26) Hexane	2.836	57	194754	8.897	ppbv	94
27) Carbon Disulfide	2.156	76	144095	9.711	ppbv #	91
28) Methyl tert-Butyl Ether	2.451	73	106845	9.804	ppbv	99
29) Chloroform	2.858	83	266332	9.040	ppbv	98
30) Cyclohexane	3.871	84	158311	8.910	ppbv	94
31) cis-1,2-Dichloroethene	2.729	61	193240	9.325	ppbv	94
32) 1,1,1-Trichloroethane	3.379	97	262261	9.177	ppbv	96
34) 2-Butanone	2.567	43	292802	10.184	ppbv	97
35) Carbon Tetrachloride	3.778	117	262129	10.300	ppbv	95
36) Benzene	3.671	78	375533	9.720	ppbv	95
37) 1,2-Dichloroethane	3.231	62	218716	10.211	ppbv	98
38) Trichloroethene	4.567	130	143714	9.110	ppbv	95
39) 1,2-Dichloropropane	4.331	63	146117	9.964	ppbv	99
40) 1,4-Dioxane	4.603	88	65661	9.809	ppbv #	96
41) Tetrahydrofuran	3.066	42	172429	10.150	ppbv	98
42) Bromodichloromethane	4.506	83	289453	10.029	ppbv	100
43) Methyl Methacrylate	4.904	69	151059	10.012	ppbv	95
44) 2,2,4-Trimethylpentane	4.658	57	684336	10.036	ppbv	99
45) t-1,3-Dichloropropene	6.441	75	147062	9.811	ppbv	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022625\
 Data File : VL042054.D
 Acq On : 26 Feb 2025 11:07
 Operator : SY/MD
 Sample : VL0226ABS01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0226ABS01

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/27/2025
 Supervised By :Mahesh Dadoda 02/27/2025

Quant Time: Feb 27 00:26:21 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Feb 26 01:08:36 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.626	75	191420	9.868	ppbv	98
47) 1,1,2-Trichloroethane	6.603	97	139001	9.337	ppbv	97
48) Dibromochloromethane	7.406	129	228694	9.808	ppbv	100
49) Bromoform	9.497	173	194002	9.581	ppbv	99
50) 4-Methyl-2-Pentanone	5.775	43	407870	10.969	ppbv	98
51) 2-Hexanone	7.454	43	327784	12.808	ppbv #	99
52) Tetrachloroethene	8.241	164	123860	9.365	ppbv	95
53) Toluene	6.953	91	418086	9.683	ppbv	98
54) 1,2-Dibromoethane	7.671	107	190454	9.462	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.940	131	171980	9.542	ppbv	100
57) Chlorobenzene	8.937	112	313487	9.482	ppbv	98
58) Ethyl Benzene	9.370	91	575798	9.805	ppbv	99
59) m/p-Xylene	9.565	91	909392m	19.459	ppbv	
60) o-Xylene	9.976	91	450481	9.588	ppbv	100
61) Styrene	9.882	104	218205	9.879	ppbv	100
62) Isopropylbenzene	10.552	105	678100	9.637	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.976	83	295251	9.543	ppbv	99
64) n-propylbenzene	10.999	120	178544	9.674	ppbv	95
65) tert-Butylbenzene	11.542	119	599970	9.369	ppbv	98
66) Benzyl Chloride	11.630	91	44022	9.009	ppbv	97
67) sec-Butylbenzene	11.779	105	863022	9.423	ppbv	99
69) p-Isopropyltoluene	11.937	119	715231	9.816	ppbv	99
70) n-Butylbenzene	12.293	91	744799	10.014	ppbv	99
71) 2-Chlorotoluene	10.914	91	520897	9.603	ppbv	99
72) 4-Ethyltoluene	11.138	105	558376	9.734	ppbv	99
73) 1,3,5-Trimethylbenzene	11.215	105	475566	9.557	ppbv	98
74) 1,2,4-Trimethylbenzene	11.549	105	520419	9.288	ppbv	99
75) 1,3-Dichlorobenzene	11.620	146	302808	9.390	ppbv	98
76) 1,4-Dichlorobenzene	11.685	146	298128	9.311	ppbv	99
77) 1,2-Dichlorobenzene	11.960	146	293043	9.364	ppbv	99
78) Hexachloro-1,3-Butadiene	13.892	225	216968	8.911	ppbv	99
79) Naphthalene	13.523	128	476359	12.399	ppbv	99
80) Naphthalene,2-methyl-	14.468	142	187061	17.792	ppbv	98
81) 1,2,4-Trichlorobenzene	13.452	180	218258	10.194	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022625\
Data File : VL042054.D
Acq On : 26 Feb 2025 11:07
Operator : SY/MD
Sample : VL0226ABS01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0226ABS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/27/2025
Supervised By :Mahesh Dadoda 02/27/2025

Quant Time: Feb 27 00:26:21 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug

QLast Update : Wed Feb 26 01:08:36 2025

Response via : Initial Calibration

