

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042179.D
 Acq On : 27 Mar 2025 12:23
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICSVVL032725

Quant Time: Mar 28 02:05:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	1.476	1.409	4.5	94	0.00
3	Chlorodifluoromethane	1.508	1.417	6.0	91	0.00
4	Chloromethane	0.603	0.565	6.3	92	0.00
5 T	Vinyl Chloride	0.639	0.548	14.2	93	0.00
6 T	Bromomethane	0.273	0.241	11.7	87	0.00
7	Chloroethane	0.242	0.215	11.2	88	0.00
8 T	Dichlorotetrafluoroethane	1.257	1.226	2.5	94	0.00
9 T	Propene	0.691	0.604	12.6	85	0.00
10 T	Heptane	1.839	1.714	6.8	89	0.00
11 T	Trichlorofluoromethane	1.437	1.358	5.5	95	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.181	1.113	5.8	93	0.00
13	Ethanol	0.065	0.033#	49.2#	108	0.00
14 T	Bromoethene	0.452	0.422	6.6	94	0.00
15 T	Acetone	1.399	1.137	18.7	95	0.00
16 T	1,3-Butadiene	0.683	0.561	17.9	91	0.00
17	tert-Butyl alcohol	1.742	1.584	9.1	91	0.00
18 T	1,1-Dichloroethene	0.552	0.500	9.4	92	0.00
19 T	Isopropyl Alcohol	0.789	0.704	10.8	93	0.00
20 T	Methylene Chloride	0.521	0.418	19.8	89	0.00
21 T	Allyl Chloride	0.986	0.903	8.4	93	0.00
22 T	trans-1,2-Dichloroethene	0.638	0.580	9.1	92	0.00
23 T	Vinyl Acetate	0.605	0.518	14.4	97	0.00
24 T	1,1-Dichloroethane	1.183	1.135	4.1	94	0.00
25 T	Ethyl Acetate	3.630	3.399	6.4	92	0.00
26 T	Hexane	1.402	1.326	5.4	93	0.00
27 T	Carbon Disulfide	1.503	1.427	5.1	93	0.00
28 T	Methyl tert-Butyl Ether	0.873	0.840	3.8	94	0.00
29 T	Chloroform	2.052	1.974	3.8	95	0.00
30 T	Cyclohexane	1.217	1.135	6.7	91	0.00
31 T	cis-1,2-Dichloroethene	1.423	1.345	5.5	91	0.00
32 T	1,1,1-Trichloroethane	2.349	2.096	10.8	95	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
34 T	2-Butanone	0.725	0.665	8.3	91	0.00
35 T	Carbon Tetrachloride	0.781	0.730	6.5	95	0.00
36 T	Benzene	0.991	0.926	6.6	91	0.00
37 T	1,2-Dichloroethane	0.568	0.549	3.3	93	0.00
38 T	Trichloroethene	0.414	0.385	7.0	94	0.00
39 T	1,2-Dichloropropane	0.361	0.343	5.0	93	0.00
40 T	1,4-Dioxane	0.192	0.162	15.6	87	0.00
41 T	Tetrahydrofuran	0.418	0.394	5.7	90	0.00
42 T	Bromodichloromethane	0.782	0.771	1.4	92	0.00
43	Methyl Methacrylate	0.420	0.392	6.7	89	0.00
44 T	2,2,4-Trimethylpentane	1.675	1.581	5.6	90	0.00
45 T	t-1,3-Dichloropropene	0.450	0.426	5.3	89	0.00
46 T	cis-1,3-Dichloropropene	0.552	0.532	3.6	88	0.00
47 T	1,1,2-Trichloroethane	0.382	0.368	3.7	92	0.00
48 T	Dibromochloromethane	0.647	0.648	-0.2	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL032725\
 Data File : VL042179.D
 Acq On : 27 Mar 2025 12:23
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL032725

Quant Time: Mar 28 02:05:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.581	0.572	1.5	90	0.00
50 T	4-Methyl-2-Pentanone	1.149	1.005	12.5	89	0.00
51 T	2-Hexanone	0.944	0.825	12.6	89	0.00
52 T	Tetrachloroethene	0.393	0.348	11.5	93	0.00
53 T	Toluene	1.175	1.118	4.9	91	0.00
54 T	1,2-Dibromoethane	0.532	0.516	3.0	93	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
56	1,1,1,2-Tetrachloroethane	0.550	0.527	4.2	94	0.00
57 T	Chlorobenzene	0.949	0.893	5.9	93	0.00
58 T	Ethyl Benzene	1.776	1.672	5.9	91	0.00
59 T	m/p-Xylene	1.418	1.331	6.1	92	0.00
60 T	o-Xylene	1.422	1.337	6.0	92	0.00
61 T	Styrene	0.703	0.667	5.1	89	0.00
62	Isopropylbenzene	2.126	1.996	6.1	93	0.00
63 T	1,1,2,2-Tetrachloroethane	0.864	0.808	6.5	93	0.00
64	n-propylbenzene	0.547	0.517	5.5	94	0.00
65	tert-Butylbenzene	1.966	1.783	9.3	94	0.00
66 T	Benzyl Chloride	0.209	0.183	12.4	82	0.00
67	sec-Butylbenzene	2.733	2.477	9.4	94	0.00
68 S	1-Bromo-4-Fluorobenzene	0.821	0.803	2.2	90	0.00
69	p-Isopropyltoluene	2.354	2.154	8.5	94	0.00
70	n-Butylbenzene	2.407	2.194	8.8	94	0.00
71	2-Chlorotoluene	1.650	1.548	6.2	92	0.00
72 T	4-Ethyltoluene	1.770	1.659	6.3	93	0.00
73 T	1,3,5-Trimethylbenzene	1.546	1.423	8.0	93	0.00
74 T	1,2,4-Trimethylbenzene	1.781	1.540	13.5	93	0.00
75 T	1,3-Dichlorobenzene	0.977	0.901	7.8	93	0.00
76 T	1,4-Dichlorobenzene	0.972	0.901	7.3	92	0.00
77 T	1,2-Dichlorobenzene	0.951	0.871	8.4	93	0.00
78 T	Hexachloro-1,3-Butadiene	0.968	0.753	22.2	86	0.00
79 T	Naphthalene	1.948	1.634	16.1	82	0.00
80 T	Naphthalene,2-methyl-	0.955	0.683	28.5	63	0.00
81 T	1,2,4-Trichlorobenzene	0.899	0.769	14.5	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0