

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WOOD06

Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG No.: Q1665

Instrument ID: MSVOA_L Calibration Date/Time: 04/15/2025 11:30

Lab File ID: VL042326.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.476	1.245		-15.65	30
Chloromethane	0.603	0.520		-13.77	30
Vinyl Chloride	0.639	0.483		-24.41	30
Bromomethane	0.273	0.211		-22.71	30
Chloroethane	0.242	0.187		-22.73	30
Tetrahydrofuran	0.418	0.397		-5.02	30
Trichlorofluoromethane	1.437	1.175		-18.23	30
1,1,2-Trichlorotrifluoroethane	1.181	0.912		-22.78	30
Dichlorotetrafluoroethane	1.257	1.062		-15.51	30
tert-Butyl alcohol	1.742	1.218		-30.08	30
Heptane	1.839	1.607		-12.62	30
1,1-Dichloroethene	0.552	0.420		-23.91	30
Acetone	1.399	1.065		-23.87	30
Carbon Disulfide	1.503	1.135		-24.48	30
Methyl tert-Butyl Ether	0.873	0.675		-22.68	30
Methylene Chloride	0.521	0.350		-32.82	30
trans-1,2-Dichloroethene	0.638	0.474		-25.7	30
1,1-Dichloroethane	1.183	0.983		-16.91	30
Cyclohexane	1.217	1.006		-17.34	30
2-Butanone	0.725	0.694		-4.28	30
Carbon Tetrachloride	0.781	0.695		-11.01	30
cis-1,2-Dichloroethene	1.423	1.208		-15.11	30
Chloroform	2.052	1.824		-11.11	30
1,1,1-Trichloroethane	2.349	1.810		-22.95	30
2,2,4-Trimethylpentane	1.675	1.593		-4.9	30
Benzene	0.991	0.928		-6.36	30
1,2-Dichloroethane	0.568	0.524		-7.75	30
Trichloroethene	0.414	0.377		-8.94	30
1,2-Dichloropropane	0.361	0.332		-8.03	30
Bromodichloromethane	0.782	0.729		-6.78	30
4-Methyl-2-Pentanone	1.149	0.969		-15.67	30
Toluene	1.175	1.010		-14.04	30
t-1,3-Dichloropropene	0.450	0.330		-26.67	30
cis-1,3-Dichloropropene	0.552	0.447		-19.02	30
1,1,2-Trichloroethane	0.382	0.357		-6.55	30
Dibromochloromethane	0.647	0.596		-7.88	30
1,2-Dibromoethane	0.532	0.481		-9.59	30
Tetrachloroethene	0.393	0.316		-19.59	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WOOD06

Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG No.: Q1665

Instrument ID: MSVOA_L Calibration Date/Time: 04/15/2025 11:30

Lab File ID: VL042326.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Chlorobenzene	0.949	0.912		-3.9	30
Ethyl Benzene	1.776	1.555		-12.44	30
m/p-Xylene	1.418	1.292		-8.89	30
o-Xylene	1.422	1.285		-9.63	30
Styrene	0.703	0.596		-15.22	30
Bromoform	0.581	0.522		-10.15	30
1,1,2,2-Tetrachloroethane	0.864	0.835		-3.36	30
2-Chlorotoluene	1.650	1.457		-11.7	30
1,3,5-Trimethylbenzene	1.546	1.355		-12.35	30
1,2,4-Trimethylbenzene	1.781	1.515		-14.94	30
1,3-Dichlorobenzene	0.977	0.927		-5.12	30
1,4-Dichlorobenzene	0.972	0.924		-4.94	30
1,2-Dichlorobenzene	0.951	0.923		-2.94	30
1,2,4-Trichlorobenzene	0.899	0.819		-8.9	30
Hexachloro-1,3-Butadiene	0.968	0.758		-21.69	30
1,3-Butadiene	0.683	0.528		-22.69	30
Naphthalene	1.948	1.849		-5.08	30
4-Ethyltoluene	1.770	1.598		-9.72	30
1-Bromo-4-Fluorobenzene	0.821	0.792		-3.53	30
Hexane	1.402	1.241		-11.48	30
Allyl Chloride	0.986	0.772		-21.7	30
1,4-Dioxane	191.626	157.635		-17.74	30
Methyl Methacrylate	0.420	0.340		-19.05	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.