

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042326.D
 Acq On : 15 Apr 2025 11:30
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 16 01:29:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	92	0.02
2 T	Dichlorodifluoromethane	1.476	1.245	15.7	82	0.00
3	Chlorodifluoromethane	1.508	1.366	9.4	86	0.00
4	Chloromethane	0.603	0.520	13.8	84	0.00
5 T	Vinyl Chloride	0.639	0.483	24.4	81	0.01
6 T	Bromomethane	0.273	0.211	22.7	76	0.01
7	Chloroethane	0.242	0.187	22.7	76	0.01
8 T	Dichlorotetrafluoroethane	1.257	1.062	15.5	81	0.01
9 T	Propene	0.691	0.587	15.1	82	0.00
10 T	Heptane	1.839	1.607	12.6	83	0.05
11 T	Trichlorofluoromethane	1.437	1.175	18.2	81	0.01
12 T	1,1,2-Trichlorotrifluoroeth	1.181	0.912	22.8	75	0.02
13	Ethanol	0.065	0.026#	60.0#	83	0.01
14 T	Bromoethene	0.452	0.346	23.5	76	0.02
15 T	Acetone	1.399	1.065	23.9	88	0.01
16 T	1,3-Butadiene	0.683	0.528	22.7	85	0.01
17	tert-Butyl alcohol	1.742	1.218	30.1#	69	0.02
18 T	1,1-Dichloroethene	0.552	0.420	23.9	77	0.02
19 T	Isopropyl Alcohol	0.789	0.606	23.2	79	0.01
20 T	Methylene Chloride	0.521	0.350	32.8#	74	0.02
21 T	Allyl Chloride	0.986	0.772	21.7	79	0.02
22 T	trans-1,2-Dichloroethene	0.638	0.474	25.7	74	0.02
23 T	Vinyl Acetate	0.605	0.419	30.7#	78	0.02
24 T	1,1-Dichloroethane	1.183	0.983	16.9	81	0.02
25 T	Ethyl Acetate	3.630	3.278	9.7	87	0.02
26 T	Hexane	1.402	1.241	11.5	86	0.02
27 T	Carbon Disulfide	1.503	1.135	24.5	73	0.02
28 T	Methyl tert-Butyl Ether	0.873	0.675	22.7	75	0.02
29 T	Chloroform	2.052	1.824	11.1	87	0.02
30 T	Cyclohexane	1.217	1.006	17.3	80	0.03
31 T	cis-1,2-Dichloroethene	1.423	1.208	15.1	81	0.02
32 T	1,1,1-Trichloroethane	2.349	1.810	22.9	81	0.03
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.04
34 T	2-Butanone	0.725	0.694	4.3	87	0.02
35 T	Carbon Tetrachloride	0.781	0.695	11.0	83	0.03
36 T	Benzene	0.991	0.928	6.4	83	0.03
37 T	1,2-Dichloroethane	0.568	0.524	7.7	82	0.03
38 T	Trichloroethene	0.414	0.377	8.9	84	0.04
39 T	1,2-Dichloropropane	0.361	0.332	8.0	83	0.04
40 T	1,4-Dioxane	0.192	0.158	17.7	78	0.05
41 T	Tetrahydrofuran	0.418	0.397	5.0	84	0.03
42 T	Bromodichloromethane	0.782	0.729	6.8	80	0.04
43	Methyl Methacrylate	0.420	0.340	19.0	71	0.05
44 T	2,2,4-Trimethylpentane	1.675	1.593	4.9	83	0.04
45 T	t-1,3-Dichloropropene	0.450	0.330	26.7	63	0.05
46 T	cis-1,3-Dichloropropene	0.552	0.447	19.0	68	0.05
47 T	1,1,2-Trichloroethane	0.382	0.357	6.5	82	0.05
48 T	Dibromochloromethane	0.647	0.596	7.9	79	0.05

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL041525\
 Data File : VL042326.D
 Acq On : 15 Apr 2025 11:30
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Apr 16 01:29:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL032725AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Mar 28 02:00:52 2025
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.581	0.522	10.2	75	0.04
50 T	4-Methyl-2-Pentanone	1.149	0.969	15.7	79	0.05
51 T	2-Hexanone	0.944	0.773	18.1	77	0.04
52 T	Tetrachloroethene	0.393	0.316	19.6	77	0.04
53 T	Toluene	1.175	1.010	14.0	75	0.05
54 T	1,2-Dibromoethane	0.532	0.481	9.6	80	0.04
55 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.04
56	1,1,1,2-Tetrachloroethane	0.550	0.520	5.5	81	0.04
57 T	Chlorobenzene	0.949	0.912	3.9	83	0.04
58 T	Ethyl Benzene	1.776	1.555	12.4	74	0.03
59 T	m/p-Xylene	1.418	1.292	8.9	77	0.03
60 T	o-Xylene	1.422	1.285	9.6	77	0.03
61 T	Styrene	0.703	0.596	15.2	69	0.03
62	Isopropylbenzene	2.126	1.931	9.2	78	0.03
63 T	1,1,2,2-Tetrachloroethane	0.864	0.835	3.4	83	0.03
64	n-propylbenzene	0.547	0.517	5.5	82	0.03
65	tert-Butylbenzene	1.966	1.752	10.9	80	0.03
66 T	Benzyl Chloride	0.209	0.132	36.8#	52	0.03
67	sec-Butylbenzene	2.733	2.494	8.7	82	0.03
68 S	1-Bromo-4-Fluorobenzene	0.821	0.792	3.5	77	0.03
69	p-Isopropyltoluene	2.354	2.103	10.7	80	0.02
70	n-Butylbenzene	2.407	2.149	10.7	80	0.03
71	2-Chlorotoluene	1.650	1.457	11.7	75	0.03
72 T	4-Ethyltoluene	1.770	1.598	9.7	78	0.03
73 T	1,3,5-Trimethylbenzene	1.546	1.355	12.4	77	0.03
74 T	1,2,4-Trimethylbenzene	1.781	1.515	14.9	79	0.02
75 T	1,3-Dichlorobenzene	0.977	0.927	5.1	83	0.03
76 T	1,4-Dichlorobenzene	0.972	0.924	4.9	82	0.03
77 T	1,2-Dichlorobenzene	0.951	0.923	2.9	86	0.03
78 T	Hexachloro-1,3-Butadiene	0.968	0.758	21.7	75	0.02
79 T	Naphthalene	1.948	1.849	5.1	81	0.02
80 T	Naphthalene,2-methyl-	0.955	0.888	7.0	71	0.02
81 T	1,2,4-Trichlorobenzene	0.899	0.819	8.9	75	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0