

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG NO.: Q1665
 Lab File ID: VL042286.D Date Analyzed: 04/09/2025
 Instrument ID: MSVOA_L Time Analyzed: 10:20
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	155631	2.78	397176	3.96	350140	8.88
UPPER LIMIT	217883	3.11	556046	4.29	490196	9.21
LOWER LIMIT	93378.6	2.45	238306	3.63	210084	8.55
EPA SAMPLE NO.						
AA-DUP-1-3242025	159221	2.78	408039	3.95	345931	8.88
AA-DUP-1-3242025DUP	156655	2.78	397453	3.95	332015	8.88
AA-1-3242025	157856	2.78	383411	3.96	331322	8.88
IA-B9-1-3242025	162172	2.78	394727	3.95	338809	8.88
IA-DUP-2-3242025	159184	2.79	384747	3.96	314874	8.88
IA-B9-2-3242025	157163	2.79	385586	3.96	326865	8.88
IA-B9-3-3242025	161865	2.79	384782	3.96	329132	8.88
VL0409ABL01	156784	2.78	391570	3.96	334054	8.88
VL0409ABS02	160536	2.78	421845	3.95	368487	8.88

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.