

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG NO.: Q1665
 Lab File ID: VL042286.D Date Analyzed: 04/09/2025
 Instrument ID: MSVOA_L Time Analyzed: 10:20
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
AA-DUP-1-3242025	0	0.00				
AA-DUP-1-3242025DUP	0	0.00				
AA-1-3242025	0	0.00				
IA-B9-1-3242025	0	0.00				
IA-DUP-2-3242025	0	0.00				
IA-B9-2-3242025	0	0.00				
IA-B9-3-3242025	0	0.00				
VL0409ABL01	0	0.00				
VL0409ABS02	0	0.00				

IS4 =

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.