

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WOOD06

Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG No.: Q1665

Instrument ID: MSVOA_L Calibration Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Calibration Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042172.D		RRF002 = VL042173.D		RRF001 = VL042174.D		
		RRF0.5 = VL042175.D		RRF0.1 = VL042176.D		RRF.03 = VL042177.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Dichlorodifluoromethane	1.404	1.483	1.514	1.602			1.476	6
Ethanol	0.029	0.073	0.086	0.110			0.065	55.5
Chloromethane	0.573	0.618	0.597	0.660			0.603	6.2
Vinyl Chloride	0.549	0.584	0.576	0.635	0.705	0.882	0.639	18.8
Bromomethane	0.256	0.266	0.268	0.327			0.273	11.2
Chloroethane	0.227	0.223	0.248	0.286			0.242	10.9
Tetrahydrofuran	0.398	0.432	0.425	0.439			0.418	5
Trichlorofluoromethane	1.339	1.480	1.457	1.562			1.437	6.6
1,1,2-Trichlorotrifluoroethane	1.120	1.188	1.215	1.247			1.181	4.5
Dichlorotetrafluoroethane	1.214	1.274	1.283	1.306			1.257	3.4
Bromoethene	0.417	0.460	0.447	0.529			0.452	10.5
tert-Butyl alcohol	1.618	1.756	1.799	1.935			1.742	7.9
Heptane	1.792	1.897	1.940	1.847			1.839	4.7
Isopropyl Alcohol	0.702	0.738	0.800	1.016			0.789	17
1,1-Dichloroethene	0.504	0.560	0.565	0.624			0.552	9
Acetone	1.113	1.469	1.573	1.699			1.399	18.7
Carbon Disulfide	1.436	1.539	1.512	1.608			1.503	5.1
Methyl tert-Butyl Ether	0.831	0.894	0.886	0.921			0.873	4.6
Methylene Chloride	0.438	0.545	0.563	0.624			0.521	15.9
trans-1,2-Dichloroethene	0.590	0.621	0.701	0.704			0.638	9.6
1,1-Dichloroethane	1.122	1.187	1.208	1.247			1.183	4.2
Cyclohexane	1.162	1.211	1.238	1.337			1.217	6.3
2-Butanone	0.667	0.770	0.771	0.754			0.725	7.6
Carbon Tetrachloride	0.703	0.748	0.732	0.709	0.782	1.069	0.781	16.6
cis-1,2-Dichloroethene	1.371	1.426	1.460	1.507			1.423	4.4
Chloroform	1.932	2.071	2.068	2.233			2.052	5.8
1,1,1-Trichloroethane	2.062	2.152	2.201	2.279	2.495	3.151	2.349	16.2
2,2,4-Trimethylpentane	1.608	1.723	1.689	1.781			1.675	5
Benzene	0.934	1.013	1.027	1.050			0.991	5.5
1,2-Dichloroethane	0.537	0.563	0.569	0.613			0.568	4.9

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WOOD06
Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG No.: Q1665
Instrument ID: MSVOA_L Calibration Date(s): 03/27/2025 03/27/2025
Heated Purge: (Y/N) N Calibration Time(s): 08:26 11:36
GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042172.D		RRF002 = VL042173.D		RRF001 = VL042174.D		
		RRF0.5 = VL042175.D		RRF0.1 = VL042176.D		RRF.03 = VL042177.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	RRF	% RSD
Trichloroethene	0.374	0.398	0.408	0.438	0.437	0.463	0.414	8
1,2-Dichloropropane	0.337	0.365	0.373	0.386			0.361	5.7
Bromodichloromethane	0.769	0.802	0.779	0.778			0.782	1.6
4-Methyl-2-Pentanone	1.027	1.211	1.214	1.286			1.149	10.8
Toluene	1.125	1.202	1.203	1.239			1.175	4.8
t-1,3-Dichloropropene	0.439	0.453	0.447	0.474			0.450	3.2
cis-1,3-Dichloropropene	0.550	0.566	0.560	0.549			0.552	2.1
1,1,2-Trichloroethane	0.364	0.379	0.382	0.419			0.382	5.9
Dibromochloromethane	0.636	0.663	0.641	0.651			0.647	1.6
1,2-Dibromoethane	0.508	0.520	0.538	0.543	0.574		0.532	4.7
Tetrachloroethene	0.343	0.371	0.361	0.401	0.359	0.572	0.393	20.6
Chlorobenzene	0.875	0.994	0.979	1.024			0.949	7.5
Ethyl Benzene	1.671	1.894	1.822	1.850			1.776	6.3
m/p-Xylene	1.321	1.484	1.472	1.505			1.418	6.7
o-Xylene	1.328	1.499	1.475	1.521			1.422	7.5
Styrene	0.683	0.735	0.722	0.709			0.703	3.9
Bromoform	0.582	0.597	0.588	0.566			0.581	2.2
1,1,2,2-Tetrachloroethane	0.796	0.893	0.889	0.907	0.800	0.986	0.864	8.8
2-Chlorotoluene	1.538	1.759	1.728	1.717			1.650	7.1
1,3,5-Trimethylbenzene	1.387	1.677	1.671	1.616			1.546	9.7
1,2,4-Trimethylbenzene	1.510	1.966	1.958	1.994			1.781	14.7
1,3-Dichlorobenzene	0.882	1.063	1.022	1.036			0.977	9
1,4-Dichlorobenzene	0.888	1.040	1.017	1.030			0.972	8
1,2-Dichlorobenzene	0.850	1.014	1.031	1.012			0.951	9.8
1,2,4-Trichlorobenzene	0.859	1.028	0.941	0.821			0.899	9.5
Hexachloro-1,3-Butadiene	0.802	1.086	1.113	1.057			0.968	16.7
1,3-Butadiene	0.573	0.608	0.705	0.952			0.683	23.4
Naphthalene	1.806	2.314	2.181	1.979	1.637		1.948	13.3
4-Ethyltoluene	1.629	1.929	1.859	1.834			1.770	8.3
1-Bromo-4-Fluorobenzene	0.813	0.828	0.810	0.832	0.827	0.820	0.821	1

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG No.: Q1665
 Instrument ID: MSVOA_L Calibration Date(s): 03/27/2025 03/27/2025
 Heated Purge: (Y/N) N Calibration Time(s): 08:26 11:36
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042172.D		RRF002 = VL042173.D		RRF001 = VL042174.D		
		RRF0.5 = VL042175.D		RRF0.1 = VL042176.D		RRF.03 = VL042177.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF.03	<u>RRF</u>	% RSD
Hexane	1.333	1.383	1.462	1.507			1.402	5.7
Allyl Chloride	0.902	1.020	1.022	1.067			0.986	7.2
1,4-Dioxane	169.386	191.108	197.154	233.701			191.626	14.1
Methyl Methacrylate	0.403	0.408	0.451	0.442			0.420	6

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.