

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WOOD06

Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG No.: Q1665

Instrument ID: MSVOA_L Calibration Date/Time: 04/09/2025 10:20

Lab File ID: VL042286.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.476	1.444		-2.17	30
Ethanol	0.065	0.030		-53.85	30
Chloromethane	0.603	0.607		0.66	30
Vinyl Chloride	0.639	0.560		-12.36	30
Bromomethane	0.273	0.250		-8.43	30
Chloroethane	0.242	0.225		-7.03	30
Tetrahydrofuran	0.418	0.441		5.5	30
Trichlorofluoromethane	1.437	1.405		-2.23	30
1,1,2-Trichlorotrifluoroethane	1.181	1.149		-2.71	30
Dichlorotetrafluoroethane	1.257	1.243		-1.11	30
Bromoethene	0.452	0.393		-13.05	30
tert-Butyl alcohol	1.742	1.415		-18.77	30
Heptane	1.839	1.800		-2.12	30
Isopropyl Alcohol	0.789	0.664		-15.84	30
1,1-Dichloroethene	0.552	0.499		-9.6	30
Acetone	1.399	1.260		-9.94	30
Carbon Disulfide	1.503	1.391		-7.45	30
Methyl tert-Butyl Ether	0.873	0.705		-19.24	30
Methylene Chloride	0.521	0.441		-15.35	30
trans-1,2-Dichloroethene	0.638	0.568		-10.97	30
1,1-Dichloroethane	1.183	1.159		-2.03	30
Cyclohexane	1.217	1.095		-10.02	30
2-Butanone	0.725	0.789		8.83	30
Carbon Tetrachloride	0.781	0.879		12.55	30
cis-1,2-Dichloroethene	1.423	1.297		-8.85	30
Chloroform	2.052	2.190		6.72	30
1,1,1-Trichloroethane	2.349	2.135		-9.11	30
2,2,4-Trimethylpentane	1.675	1.869		11.58	30
Benzene	0.991	1.071		8.07	30
1,2-Dichloroethane	0.568	0.647		13.91	30
Trichloroethene	0.414	0.429		3.62	30
1,2-Dichloropropane	0.361	0.387		7.2	30
Bromodichloromethane	0.782	0.904		15.6	30
4-Methyl-2-Pentanone	1.149	1.122		-2.35	30
Toluene	1.175	1.171		-0.34	30
t-1,3-Dichloropropene	0.450	0.371		-17.56	30
cis-1,3-Dichloropropene	0.552	0.493		-10.69	30
1,1,2-Trichloroethane	0.382	0.434		13.61	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WOOD06

Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG No.: Q1665

Instrument ID: MSVOA_L Calibration Date/Time: 04/09/2025 10:20

Lab File ID: VL042286.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dibromochloromethane	0.647	0.705		8.96	30
1,2-Dibromoethane	0.532	0.578		8.65	30
Tetrachloroethene	0.393	0.357		-9.16	30
Chlorobenzene	0.949	1.140		20.13	30
Ethyl Benzene	1.776	1.924		8.33	30
m/p-Xylene	1.418	1.658		16.92	30
o-Xylene	1.422	1.659		16.67	30
Styrene	0.703	0.718		2.13	30
Bromoform	0.581	0.630		8.43	30
1,1,2,2-Tetrachloroethane	0.864	1.091		26.27	30
2-Chlorotoluene	1.650	1.869		13.27	30
1,3,5-Trimethylbenzene	1.546	1.772		14.62	30
1,2,4-Trimethylbenzene	1.781	2.025		13.7	30
1,3-Dichlorobenzene	0.977	1.218		24.67	30
1,4-Dichlorobenzene	0.972	1.210		24.49	30
1,2-Dichlorobenzene	0.951	1.197		25.87	30
1,2,4-Trichlorobenzene	0.899	1.016		13.01	30
Hexachloro-1,3-Butadiene	0.968	0.955		-1.34	30
1,3-Butadiene	0.683	0.610		-10.69	30
Naphthalene	1.948	2.309		18.53	30
4-Ethyltoluene	1.770	2.081		17.57	30
1-Bromo-4-Fluorobenzene	0.821	0.815		-0.73	30
Hexane	1.402	1.381		-1.5	30
Allyl Chloride	0.986	0.899		-8.82	30
1,4-Dioxane	191.626	182.801		-4.61	30
Methyl Methacrylate	0.420	0.390		-7.14	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.