

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q1665	Analytical Method:	SWTO-15
Client:	Woodard & Curran	Datafile :	VL042314.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0411ABS01	Dichlorodifluoromethane	10	9.90	ppbv	99			40 (70)	160 (130)	
	Ethanol	10	4.20	ppbv	42		*	70 (70)	130 (130)	
	Chloromethane	10	10.2	ppbv	102			40 (70)	160 (130)	
	Vinyl Chloride	10	8.70	ppbv	87			70 (70)	130 (130)	
	Bromomethane	10	8.90	ppbv	89			40 (70)	160 (130)	
	Chloroethane	10	9.30	ppbv	93			40 (70)	160 (130)	
	Tetrahydrofuran	10	10.9	ppbv	109			70 (70)	130 (130)	
	Trichlorofluoromethane	10	9.60	ppbv	96			40 (70)	160 (130)	
	1,1,2-Trichlorotrifluoroethane	10	9.10	ppbv	91			70 (70)	130 (130)	
	Dichlorotetrafluoroethane	10	9.80	ppbv	98			70 (70)	130 (130)	
	Bromoethene	10	8.80	ppbv	88			70 (70)	130 (130)	
	tert-Butyl Alcohol	10	7.80	ppbv	78			70 (70)	130 (130)	
	Heptane	10	9.90	ppbv	99			70 (70)	130 (130)	
	Isopropyl Alcohol	10	8.40	ppbv	84			70 (70)	130 (130)	
	1,1-Dichloroethene	10	8.40	ppbv	84			70 (70)	130 (130)	
	Acetone	10	9.20	ppbv	92			40 (70)	160 (130)	
	Carbon disulfide	10	8.80	ppbv	88			40 (70)	160 (130)	
	Methyl tert-butyl Ether	10	7.90	ppbv	79			70 (70)	130 (130)	
	Methylene Chloride	10	8.10	ppbv	81			70 (70)	130 (130)	
	trans-1,2-Dichloroethene	10	8.70	ppbv	87			70 (70)	130 (130)	
	1,1-Dichloroethane	10	9.80	ppbv	98			70 (70)	130 (130)	
	Cyclohexane	10	9.00	ppbv	90			70 (70)	130 (130)	
	2-Butanone	10	11.3	ppbv	113			40 (70)	160 (130)	
	Carbon Tetrachloride	10	11.2	ppbv	112			70 (70)	130 (130)	
	cis-1,2-Dichloroethene	10	9.20	ppbv	92			70 (70)	130 (130)	
	Chloroform	10	10.5	ppbv	105			70 (70)	130 (130)	
	1,1,1-Trichloroethane	10	8.80	ppbv	88			70 (70)	130 (130)	
	2,2,4-Trimethylpentane	10	11.4	ppbv	114			70 (70)	130 (130)	
	Benzene	10	11.1	ppbv	111			70 (70)	130 (130)	
	1,2-Dichloroethane	10	11.5	ppbv	115			70 (70)	130 (130)	
	Trichloroethene	10	10.6	ppbv	106			70 (70)	130 (130)	
	1,2-Dichloropropane	10	11.2	ppbv	112			70 (70)	130 (130)	
	Bromodichloromethane	10	11.7	ppbv	117			70 (70)	130 (130)	
	4-Methyl-2-Pentanone	10	10.1	ppbv	101			40 (70)	160 (130)	
	Toluene	10	10.1	ppbv	101			70 (70)	130 (130)	
	t-1,3-Dichloropropene	10	8.30	ppbv	83			70 (70)	130 (130)	
	cis-1,3-Dichloropropene	10	9.20	ppbv	92			70 (70)	130 (130)	
	1,1,2-Trichloroethane	10	11.7	ppbv	117			70 (70)	130 (130)	
	Dibromochloromethane	10	11.2	ppbv	112			70 (70)	130 (130)	
	1,2-Dibromoethane	10	11.2	ppbv	112			70 (70)	130 (130)	
	Tetrachloroethene	10	9.50	ppbv	95			70 (70)	130 (130)	
	Chlorobenzene	10	12.5	ppbv	125			70 (70)	130 (130)	
	Ethyl Benzene	10	11.1	ppbv	111			70 (70)	130 (130)	
	m/p-Xylene	20	24.2	ppbv	121			70 (70)	130 (130)	
	o-Xylene	10	11.9	ppbv	119			70 (70)	130 (130)	
	Styrene	10	10.5	ppbv	105			70 (70)	130 (130)	
	Bromoform	10	11.3	ppbv	113			70 (70)	130 (130)	
	1,1,2,2-Tetrachloroethane	10	13.2	ppbv	132		*	70 (70)	130 (130)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q1665</u>	Analytical Method:	<u>SWTO-15</u>
Client:	<u>Woodard & Curran</u>	Datafile :	<u>VL042314.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0411ABS01	2-Chlorotoluene	10	11.8	ppbv	118			70 (70)	130 (130)	
	1,3,5-Trimethylbenzene	10	11.9	ppbv	119			70 (70)	130 (130)	
	1,2,4-Trimethylbenzene	10	11.9	ppbv	119			70 (70)	130 (130)	
	1,3-Dichlorobenzene	10	12.8	ppbv	128			70 (70)	130 (130)	
	1,4-Dichlorobenzene	10	12.9	ppbv	129			70 (70)	130 (130)	
	1,2-Dichlorobenzene	10	13.2	ppbv	132		*	70 (70)	130 (130)	
	1,2,4-Trichlorobenzene	10	10.8	ppbv	108			70 (70)	130 (130)	
	Hexachloro-1,3-butadiene	10	9.80	ppbv	98			70 (70)	130 (130)	
	Naphthalene	10	10.9	ppbv	109			40 (70)	160 (130)	
	1,3-Butadiene	10	9.10	ppbv	91			70 (70)	130 (130)	
	4-Ethyltoluene	10	12.1	ppbv	121			70 (70)	130 (130)	
	Hexane	10	10.0	ppbv	100			70 (70)	130 (130)	
	Allyl Chloride	10	8.90	ppbv	89			70 (70)	130 (130)	
	1,4-Dioxane	10	9.40	ppbv	94			40 (70)	160 (130)	
	Methyl methacrylate	10	9.50	ppbv	95			70 (70)	130 (130)	