

Duplicate Sample Summary

Lab Sample Id :	Q1665-01DUP	Q1665-01
Client Id :	AA-DUP-1-3242025DUP	AA-DUP-1-3242025
DF :	1	1
Datafile :	VL042291.D	VL042290.D
Anal Date & Time :	04/09/2025 14:02	04/09/2025 13:25

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0	0	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	0.33	0.36	8.7
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	3.6	3.6	0
Allyl Chloride	0	0	0
Benzene	0.24	0.25	4.1
Bromodichloromethane	0	0	0
Bromoethene	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0

Duplicate Sample Summary

Lab Sample Id :	Q1665-01DUP	Q1665-01
Client Id :	AA-DUP-1-3242025DUP	AA-DUP-1-3242025
DF :	1	1
Datafile :	VL042291.D	VL042290.D
Anal Date & Time :	04/09/2025 14:02	04/09/2025 13:25

Parameter	Result	Result	RPD
Carbon Tetrachloride	0.08	0.07	13.3
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.55	0.56	1.8
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.38	0.4	5.1
Dichlorotetrafluoroethane	0	0	0
Ethanol	12.8	13.4	4.6
Ethyl Benzene	0	0	0
Heptane	0.1	0.11	9.5
Hexachloro-1,3-Butadiene	0	0	0
Hexane	1.4	1.6	13.3
Isopropyl Alcohol	1.9	1.8	5.4
m/p-Xylene	0.12	0	200 *
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	0.95	0.95	0
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0	0	0
Tetrahydrofuran	0	0	0
Toluene	0.28	0.32	13.3
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0

Duplicate Sample Summary

Lab Sample Id :	Q1665-01DUP	Q1665-01
Client Id :	AA-DUP-1-3242025DUP	AA-DUP-1-3242025
DF :	1	1
Datafile :	VL042291.D	VL042290.D
Anal Date & Time :	04/09/2025 14:02	04/09/2025 13:25

Parameter	Result	Result	RPD
Trichlorofluoromethane	0.22	0.21	4.7
Vinyl Chloride	0	0	0