

Duplicate Sample Summary

Lab Sample Id :	Q1665-01DUP	Q1665-01
Client Id :	AA-DUP-1-3242025DUP	AA-DUP-1-3242025
DF :	1	1
Datafile :	VL042291.D	VL042290.D
Anal Date & Time :	04/09/2025 14:02	04/09/2025 13:25

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0	0	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	0.33	0.36	8.7
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	3.6	3.6	0
Allyl Chloride	0	0	0
Benzene	0.24	0.25	4.1
Bromodichloromethane	0	0	0
Bromoethene	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0

Duplicate Sample Summary

Lab Sample Id :	Q1665-01DUP	Q1665-01
Client Id :	AA-DUP-1-3242025DUP	AA-DUP-1-3242025
DF :	1	1
Datafile :	VL042291.D	VL042290.D
Anal Date & Time :	04/09/2025 14:02	04/09/2025 13:25

Parameter	Result	Result	RPD
Carbon Tetrachloride	0.08	0.07	13.3
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.55	0.56	1.8
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.38	0.4	5.1
Dichlorotetrafluoroethane	0	0	0
Ethanol	12.8	13.4	4.6
Ethyl Benzene	0	0	0
Heptane	0.1	0.11	9.5
Hexachloro-1,3-Butadiene	0	0	0
Hexane	1.4	1.6	13.3
Isopropyl Alcohol	1.9	1.8	5.4
m/p-Xylene	0.12	0	200 *
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	0.95	0.95	0
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0	0	0
Tetrahydrofuran	0	0	0
Toluene	0.28	0.32	13.3
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0

Duplicate Sample Summary

Lab Sample Id : Q1665-01DUP Q1665-01
Client Id : AA-DUP-1-3242025DUP AA-DUP-1-3242025
DF : 1 1
Datafile : VL042291.D VL042290.D
Anal Date & Time : 04/09/2025 14:02 04/09/2025 13:25

Parameter	Result	Result	RPD
Trichlorofluoromethane	0.22	0.21	4.7
Vinyl Chloride	0	0	0

Duplicate Sample Summary

Lab Sample Id :	Q1665-03DUP	Q1665-03
Client Id :	IA-B10-2-3242025DUP	IA-B10-2-3242025
DF :	1	1
Datafile :	VL042318.D	VL042317.D
Anal Date & Time :	04/11/2025 13:52	04/11/2025 13:15

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0.13	0.12	8
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	26.6	26.8	0.75
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	82.1	82.7	0.73
Allyl Chloride	0	0	0
Benzene	0.6	0.6	0
Bromodichloromethane	0	0	0
Bromoethene	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0

Duplicate Sample Summary

Lab Sample Id :	Q1665-03DUP	Q1665-03
Client Id :	IA-B10-2-3242025DUP	IA-B10-2-3242025
DF :	1	1
Datafile :	VL042318.D	VL042317.D
Anal Date & Time :	04/11/2025 13:52	04/11/2025 13:15

Parameter	Result	Result	RPD
Carbon Tetrachloride	0.09	0.09	0
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	1.4	1.4	0
Chloromethane	0.59	0.59	0
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.42	0.4	4.9
Dichlorotetrafluoroethane	0	0	0
Ethanol	1300	1300	0
Ethyl Benzene	0	0	0
Heptane	0.19	0.17	11.1
Hexachloro-1,3-Butadiene	0	0	0
Hexane	2.7	2.5	7.7
Isopropyl Alcohol	83.2	81.1	2.6
m/p-Xylene	0.2	0	200 *
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	1.6	1.6	0
Naphthalene	0	0	0
o-Xylene	0.12	0.12	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0.46	0.47	2.2
Tetrachloroethene	0.04	0.04	0
Tetrahydrofuran	0	0	0
Toluene	0.51	0.49	4
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0

Duplicate Sample Summary

Lab Sample Id :	Q1665-03DUP	Q1665-03
Client Id :	IA-B10-2-3242025DUP	IA-B10-2-3242025
DF :	1	1
Datafile :	VL042318.D	VL042317.D
Anal Date & Time :	04/11/2025 13:52	04/11/2025 13:15

Parameter	Result	Result	RPD
Trichlorofluoromethane	0.22	0.21	4.7
Vinyl Chloride	0	0	0