

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG NO.: Q1665
 Lab File ID: VL042326.D Date Analyzed: 04/15/2025
 Instrument ID: MSVOA_L Time Analyzed: 11:30
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	172263	2.81	457227	3.99	411493	8.91
UPPER LIMIT	241168	3.14	640118	4.32	576090	9.24
LOWER LIMIT	103358	2.48	274336	3.66	246896	8.58
EPA SAMPLE NO.						
IA-B10-4-3242025DL2	164444	2.81	385228	4.00	327587	8.92
IA-B10-3-3242025DL	168977	2.81	401158	3.99	339800	8.91
VL0415ABL01	173729	2.81	447320	3.99	385318	8.92
VL0415ABS01	171986	2.81	444595	3.99	399604	8.92

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.