

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG NO.: Q1665
Lab File ID: VL042416.D Date Analyzed: 04/28/2025
Instrument ID: MSVOA_L Time Analyzed: 08:44
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	121989	2.78	345031	3.96	305240	8.88
UPPER LIMIT	170785	3.11	483043	4.29	427336	9.21
LOWER LIMIT	73193.4	2.45	207019	3.63	183144	8.55
EPA SAMPLE NO.						
IA-B2-3-3242025	141298	2.78	401917	3.95	351139	8.88
IA-DUP-1-3242025	141186	2.78	402115	3.95	352885	8.87
IA-B2-6-3242025	140992	2.78	398020	3.95	341014	8.88
IA-B2-5-3242025	143445	2.78	401346	3.95	346372	8.87
IA-B2-2-3242025	134276	2.78	394249	3.95	337638	8.88
IA-B2-1-3242025	134924	2.78	377834	3.96	330143	8.88
IA-B2-4-3242025	133015	2.78	381932	3.96	330693	8.88
VL0428ABL01	121027	2.78	346921	3.95	302157	8.87
VL0428ABS01	122448	2.78	337202	3.95	299130	8.88

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.