

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
Lab Code: CHEM Case No.: Q1665 SAS No.: Q1665 SDG NO.: Q1665
Lab File ID: VL042436.D Date Analyzed: 04/29/2025
Instrument ID: MSVOA_L Time Analyzed: 08:39
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	114553	2.78	311268	3.95	269987	8.88
UPPER LIMIT	160374	3.11	435775	4.28	377982	9.21
LOWER LIMIT	68731.8	2.45	186761	3.62	161992	8.55
EPA SAMPLE NO.						
IA-DUP-1-3242025DL	116268	2.80	307149	3.97	260772	8.90
IA-B2-2-3242025DL	114403	2.79	305602	3.97	261578	8.90
IA-B2-1-3242025DL	114227	2.80	302811	3.97	257605	8.90
VL0429ABL01	119014	2.79	328601	3.96	276498	8.89
VL0429ABS01	116772	2.79	317590	3.96	272312	8.88

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.