

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135236

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	04/01/2025	11:00
2	Q1666-01		1	15	100.00	>212.0	>212.0	04/01/2025	11:30
3	Q1666-01DUP		1	15	100.00	>212.0	>212.0	04/01/2025	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 135236

WorkList Name : fp-3-31

WorkList ID : 188640

Department : Wet-Chemistry

Date : 03-31-2025 12:27:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1666-01	TW-WTS-05	Water	Flash Point	Cool 4 deg C	ENTA05	I31	03/25/2025	1010B

Date/Time 04/01/2025 10:45
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC

Date/Time 04/01/2025 12:40
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC