

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1666

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135226

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.4	10	104	90-110	03/21/2025
Chloride	mg/L	3	3	100	90-110	03/21/2025
Fluoride	mg/L	2	2	100	90-110	03/21/2025
Nitrite	mg/L	3.1	3	103	90-110	03/21/2025
Nitrate	mg/L	2.6	2.5	104	90-110	03/21/2025
Sulfate	mg/L	15.3	15	102	90-110	03/21/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	03/21/2025
Sample ID: CCV1						
Bromide	mg/L	10.7	10	107	90-110	03/28/2025
Chloride	mg/L	3.2	3	107	90-110	03/28/2025
Fluoride	mg/L	2.1	2	105	90-110	03/28/2025
Nitrite	mg/L	3.2	3	107	90-110	03/28/2025
Nitrate	mg/L	2.6	2.5	104	90-110	03/28/2025
Sulfate	mg/L	15.5	15	103	90-110	03/28/2025
Orthophosphate as P	mg/L	5.4	5	108	90-110	03/28/2025
Sample ID: CCV2						
Bromide	mg/L	10.7	10	107	90-110	03/28/2025
Chloride	mg/L	3.2	3	107	90-110	03/28/2025
Fluoride	mg/L	2.1	2	105	90-110	03/28/2025
Nitrite	mg/L	3.2	3	107	90-110	03/28/2025
Nitrate	mg/L	2.6	2.5	104	90-110	03/28/2025
Sulfate	mg/L	15.6	15	104	90-110	03/28/2025
Orthophosphate as P	mg/L	5.4	5	108	90-110	03/28/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1666

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135236

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	82.4	81	102	78-84	04/01/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1666

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135351

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	04/09/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	04/09/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	04/09/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1666

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135351

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------