

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q1666
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1666-01
Client ID:	TW-WTS-05MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	12.4		1.00	J	10	1	114		03/28/2025
Chloride	mg/L	80-120	52.5	OR	51.4	OR	3	1	37	*	03/28/2025
Fluoride	mg/L	80-120	2.40		0.52		2	1	94		03/28/2025
Nitrite	mg/L	80-120	3.10		0.20	J	3	1	97		03/28/2025
Nitrate	mg/L	80-120	2.60		0.18	J	2.5	1	97		03/28/2025
Sulfate	mg/L	80-120	1110	OR	1150	OR	15	1	-267	*	03/28/2025
Orthophosphate as P	mg/L	80-120	5.10		0.34	U	5	1	102		03/28/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q1666
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1666-01
Client ID:	TW-WTS-05MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	12.3		1.00	J	10	1	113		03/28/2025
Chloride	mg/L	80-120	52.6	OR	51.4	OR	3	1	40	*	03/28/2025
Fluoride	mg/L	80-120	2.40		0.52		2	1	94		03/28/2025
Nitrite	mg/L	80-120	3.10		0.20	J	3	1	97		03/28/2025
Nitrate	mg/L	80-120	2.60		0.18	J	2.5	1	97		03/28/2025
Sulfate	mg/L	80-120	1110	OR	1150	OR	15	1	-267	*	03/28/2025
Orthophosphate as P	mg/L	80-120	5.20		0.34	U	5	1	104		03/28/2025