



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT SDG No.: Q1666
Contract: ENTA05 Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	101	40 - 160	CV	04/02/2025	14:19	LB135269
CRI01	Cadmium	5.72	6.0	95	40 - 160	P	04/02/2025	15:55	LB135273
	Copper	22.1	20.0	110	40 - 160	P	04/02/2025	15:55	LB135273
	Lead	11.9	12.0	99	40 - 160	P	04/02/2025	15:55	LB135273
	Nickel	38.7	40.0	97	40 - 160	P	04/02/2025	15:55	LB135273
	Zinc	41.8	40.0	104	40 - 160	P	04/02/2025	15:55	LB135273
CRI01	Cadmium	5.76	6.0	96	40 - 160	P	04/03/2025	12:20	LB135289
	Copper	22.8	20.0	114	40 - 160	P	04/03/2025	12:20	LB135289
	Lead	11.6	12.0	96	40 - 160	P	04/03/2025	12:20	LB135289
	Nickel	38.7	40.0	97	40 - 160	P	04/03/2025	12:20	LB135289
	Zinc	42.8	40.0	107	40 - 160	P	04/03/2025	12:20	LB135289