



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q1666</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1666</u>	SAS No.:	<u>Q1666</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB73	Mercury	0.20	+/-0.20	U	0.20	CV	04/02/2025	14:12	LB135269

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT **SDG No.:** Q1666
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1666 **SAS No.:** Q1666

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB23	Mercury	0.20	+/-0.20	U	0.20	CV	04/02/2025	14:17	LB135269
CCB24	Mercury	0.20	+/-0.20	U	0.20	CV	04/02/2025	14:46	LB135269
CCB25	Mercury	0.20	+/-0.20	U	0.20	CV	04/02/2025	15:05	LB135269

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1666</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1666</u>	SAS No.: <u>Q1666</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	04/02/2025	15:51	LB135273
	Copper	20.0	+/-20.0	U	20.0	P	04/02/2025	15:51	LB135273
	Lead	12.0	+/-12.0	U	12.0	P	04/02/2025	15:51	LB135273
	Nickel	40.0	+/-40.0	U	40.0	P	04/02/2025	15:51	LB135273
	Zinc	40.0	+/-40.0	U	40.0	P	04/02/2025	15:51	LB135273

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1666</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1666</u>	SAS No.: <u>Q1666</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	04/02/2025	16:43	LB135273
	Copper	20.0	+/-20.0	U	20.0	P	04/02/2025	16:43	LB135273
	Lead	12.0	+/-12.0	U	12.0	P	04/02/2025	16:43	LB135273
	Nickel	40.0	+/-40.0	U	40.0	P	04/02/2025	16:43	LB135273
	Zinc	40.0	+/-40.0	U	40.0	P	04/02/2025	16:43	LB135273
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	04/02/2025	17:39	LB135273
	Copper	20.0	+/-20.0	U	20.0	P	04/02/2025	17:39	LB135273
	Lead	12.0	+/-12.0	U	12.0	P	04/02/2025	17:39	LB135273
	Nickel	40.0	+/-40.0	U	40.0	P	04/02/2025	17:39	LB135273
	Zinc	40.0	+/-40.0	U	40.0	P	04/02/2025	17:39	LB135273
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	04/02/2025	17:58	LB135273
	Copper	20.0	+/-20.0	U	20.0	P	04/02/2025	17:58	LB135273
	Lead	12.0	+/-12.0	U	12.0	P	04/02/2025	17:58	LB135273
	Nickel	40.0	+/-40.0	U	40.0	P	04/02/2025	17:58	LB135273
	Zinc	40.0	+/-40.0	U	40.0	P	04/02/2025	17:58	LB135273

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1666</u>							
Contract: <u>ENTA05</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1666</u>		SAS No.: <u>Q1666</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	12:15	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	12:15	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	12:15	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q1666							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q1666	SAS No.: Q1666						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	12:47	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	12:47	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	12:47	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	13:57	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	13:57	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	13:57	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
CCB03	Cadmium	0.26	+/-6.00	J	6.00	P	04/03/2025	14:43	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	14:43	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	14:43	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	15:36	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	15:36	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	15:36	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
CCB05	Cadmium	0.38	+/-6.00	J	6.00	P	04/03/2025	16:40	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	16:40	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	16:40	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	18:00	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	18:00	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	18:00	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
CCB07	Cadmium	0.23	+/-6.00	J	6.00	P	04/03/2025	18:33	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	18:33	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	18:33	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT **SDG No.:** Q1666
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1666 **SAS No.:** Q1666

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------