

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client:	<u>ENTACT</u>	SDG No.:	<u>Q1666</u>
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>
ICS Source:	<u>EPA</u>	Case No.:	<u>Q1666</u>
		SAS No.:	<u>Q1666</u>
		Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-0.59	1.0	59	-5	7	04/02/2025	15:59	LB135273
	Copper	-8.99	2.0	450	-18	22	04/02/2025	15:59	LB135273
	Lead	8.38			-12	12	04/02/2025	15:59	LB135273
	Nickel	2.54	2.0	127	-38	42	04/02/2025	15:59	LB135273
	Zinc	4.72			-40	40	04/02/2025	15:59	LB135273
ICSAB01	Cadmium	1010	972	104	826	1120	04/02/2025	16:04	LB135273
	Copper	498	511	98	434	588	04/02/2025	16:04	LB135273
	Lead	56.3	49.0	115	37	61	04/02/2025	16:04	LB135273
	Nickel	1020	954	107	810	1100	04/02/2025	16:04	LB135273
	Zinc	1080	952	113	809	1095	04/02/2025	16:04	LB135273
ICSA01	Cadmium	-1.36	1.0	136	-5	7	04/03/2025	12:24	LB135289
	Copper	3.95	2.0	198	-18	22	04/03/2025	12:24	LB135289
	Lead	4.36			-12	12	04/03/2025	12:24	LB135289
	Nickel	1.10	2.0	55	-38	42	04/03/2025	12:24	LB135289
	Zinc	5.44			-40	40	04/03/2025	12:24	LB135289
ICSAB01	Cadmium	1000	972	103	826	1120	04/03/2025	12:28	LB135289
	Copper	512	511	100	434	588	04/03/2025	12:28	LB135289
	Lead	55.2	49.0	113	37	61	04/03/2025	12:28	LB135289
	Nickel	1010	954	106	810	1100	04/03/2025	12:28	LB135289
	Zinc	1070	952	112	809	1095	04/03/2025	12:28	LB135289