

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1666
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1666 **SAS No.:** Q1666

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167422BS							
Cadmium	ug/L	100	93.8		94	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Lead	ug/L	500	467		93	80 - 120	P
Nickel	ug/L	250	241		96	80 - 120	P
Zinc	ug/L	100	101		101	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1666
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1666 **SAS No.:** Q1666

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167425BS Mercury	ug/L	4.0	3.42		86	80 - 120	CV