

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666 SDG NO.: Q1666

Continuing Calib Date: 03/28/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 19:42 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.52       | 6.52      | 6.42      | 6.62  | 0.00       |
| Aroclor-1254-2       | (2) | 6.74       | 6.73      | 6.63      | 6.83  | -0.01      |
| Aroclor-1254-3       | (3) | 7.10       | 7.10      | 7.00      | 7.20  | 0.00       |
| Aroclor-1254-4       | (4) | 7.39       | 7.38      | 7.28      | 7.48  | -0.01      |
| Aroclor-1254-5       | (5) | 7.80       | 7.79      | 7.69      | 7.89  | -0.01      |
| Tetrachloro-m-xylene |     | 4.52       | 4.51      | 4.41      | 4.61  | -0.01      |
| Decachlorobiphenyl   |     | 10.25      | 10.23     | 10.13     | 10.33 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666 SDG NO.: Q1666

Continuing Calib Date: 03/28/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 19:42 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.71       | 5.71      | 5.61      | 5.81 | 0.00       |
| Aroclor-1254-2 (2)   | 5.86       | 5.86      | 5.76      | 5.96 | 0.00       |
| Aroclor-1254-3 (3)   | 6.26       | 6.26      | 6.16      | 6.36 | 0.00       |
| Aroclor-1254-4 (4)   | 6.49       | 6.49      | 6.39      | 6.59 | 0.00       |
| Aroclor-1254-5 (5)   | 6.91       | 6.91      | 6.81      | 7.01 | 0.00       |
| Tetrachloro-m-xylene | 3.82       | 3.82      | 3.72      | 3.92 | 0.00       |
| Decachlorobiphenyl   | 8.86       | 8.86      | 8.76      | 8.96 | 0.00       |