

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666 SDG NO.: Q1666

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 03/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP070994.D Time Analyzed: 23:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.672	5.563	5.763	515.870	500.000	3.2
Aroclor-1016-2	5.694	5.585	5.785	521.960	500.000	4.4
Aroclor-1016-3	5.756	5.648	5.848	511.070	500.000	2.2
Aroclor-1016-4	5.853	5.745	5.945	552.650	500.000	10.5
Aroclor-1016-5	6.147	6.038	6.238	549.950	500.000	10.0
Aroclor-1260-1	7.266	7.156	7.356	569.940	500.000	14.0
Aroclor-1260-2	7.519	7.410	7.610	573.260	500.000	14.7
Aroclor-1260-3	7.878	7.768	7.968	555.320	500.000	11.1
Aroclor-1260-4	8.103	7.992	8.192	533.840	500.000	6.8
Aroclor-1260-5	8.423	8.311	8.511	555.310	500.000	11.1
Decachlorobiphenyl	10.242	10.125	10.325	56.260	50.000	12.5
Tetrachloro-m-xylene	4.518	4.412	4.612	53.000	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666 SDG NO.: Q1666

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 03/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP070994.D Time Analyzed: 23:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.903	4.803	5.003	496.240	500.000	-0.8
Aroclor-1016-2	4.922	4.822	5.022	476.970	500.000	-4.6
Aroclor-1016-3	5.099	4.999	5.199	531.400	500.000	6.3
Aroclor-1016-4	5.141	5.041	5.241	531.860	500.000	6.4
Aroclor-1016-5	5.356	5.255	5.455	498.170	500.000	-0.4
Aroclor-1260-1	6.391	6.292	6.492	514.830	500.000	3.0
Aroclor-1260-2	6.580	6.480	6.680	528.940	500.000	5.8
Aroclor-1260-3	6.734	6.634	6.834	544.360	500.000	8.9
Aroclor-1260-4	7.205	7.105	7.305	517.810	500.000	3.6
Aroclor-1260-5	7.446	7.346	7.546	492.440	500.000	-1.5
Decachlorobiphenyl	8.860	8.759	8.959	53.720	50.000	7.4
Tetrachloro-m-xylene	3.816	3.716	3.916	48.350	50.000	-3.3