

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666 SDG No.: Q1666

Instrument ID: BNA_G Calibration Date/Time: 04/01/2025 11:38

Lab File ID: BG064130.D Init. Calib. Date(s): 03/05/2025 03/05/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:02 13:44

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.281	1.136		-11.3	
Phenol-d6	1.742	1.624		-6.8	
Phenol	1.784	1.651		-7.5	20.0
1,4-Dichlorobenzene	1.548	1.380		-10.9	20.0
Nitrobenzene-d5	0.362	0.374		3.3	
1,2,4-Trichlorobenzene	0.331	0.319		-3.6	
Naphthalene	1.078	1.053		-2.3	
2-Fluorobiphenyl	1.318	1.245		-5.5	
2,4,6-Tribromophenol	0.222	0.244		9.9	
Terphenyl-d14	0.989	0.963		-2.6	

All other compounds must meet a minimum RRF of 0.010.