

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030525\
 Data File : BG064052.D
 Acq On : 5 Mar 2025 13:44
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

SSTDICC080

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/06/2025
 Supervised By :mohammad ahmed 03/07/2025

Quant Time: Mar 05 15:25:59 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030525.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Mar 05 14:45:06 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.865	152	38973	20.000	ng	0.00
21) Naphthalene-d8	10.656	136	174562	20.000	ng	0.00
39) Acenaphthene-d10	14.493	164	119753	20.000	ng	0.00
64) Phenanthrene-d10	17.231	188	261353	20.000	ng	0.00
76) Chrysene-d12	21.467	240	257789	20.000	ng	0.00
86) Perylene-d12	24.481	264	274554	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.456	112	406682	162.936	ng	0.00
7) Phenol-d6	7.037	99	562221	165.580	ng	0.00
23) Nitrobenzene-d5	9.029	82	570757	180.687	ng	0.01
42) 2,4,6-Tribromophenol	15.979	330	230590	173.228	ng	0.00
45) 2-Fluorobiphenyl	13.118	172	1198476	151.908	ng	0.00
79) Terphenyl-d14	19.857	244	1912762	150.030	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.376	88	86434	76.411	ng	97
3) Pyridine	3.770	79	194608	70.740	ng	99
4) n-Nitrosodimethylamine	3.682	42	166554	84.738	ng	98
6) Aniline	7.196	93	260509	78.188	ng	97
8) 2-Chlorophenol	7.436	128	221296	82.552	ng	96
10) Phenol	7.066	94	287854	82.802	ng	97
11) bis(2-Chloroethyl)ether	7.295	93	217240	79.707	ng	99
12) 1,3-Dichlorobenzene	7.760	146	235126	79.873	ng	96
13) 1,4-Dichlorobenzene	7.901	146	238817	79.149	ng	96
14) 1,2-Dichlorobenzene	8.224	146	233348	80.201	ng	97
15) Benzyl Alcohol	8.106	79	227176	86.581	ng	94
16) 2,2'-oxybis(1-Chloropr...	8.400	45	502220	81.950	ng	99
17) 2-Methylphenol	8.306	107	198161	85.886	ng	98
18) Hexachloroethane	8.946	117	90416	85.647	ng	98
19) n-Nitroso-di-n-propyla...	8.682	70	197732	82.978	ng	96
20) 3+4-Methylphenols	8.641	107	269552	84.861	ng	92
22) Acetophenone	8.688	105	379097	79.207	ng	99
24) Nitrobenzene	9.064	77	285702	87.518	ng	99
25) Isophorone	9.593	82	507481	80.266	ng	96
26) 2-Nitrophenol	9.769	139	102389	79.399	ng	96
27) 2,4-Dimethylphenol	9.834	122	159995	84.413	ng	94
28) bis(2-Chloroethoxy)met...	10.074	93	310803	81.082	ng	99
29) 2,4-Dichlorophenol	10.304	162	208425	87.086	ng	98
30) 1,2,4-Trichlorobenzene	10.521	180	231572	80.151	ng	99
31) Naphthalene	10.709	128	752327	79.925	ng	99
32) Benzoic acid	10.022	122	153536m	81.016	ng	
33) 4-Chloroaniline	10.815	127	281930	81.948	ng	100
34) Hexachlorobutadiene	10.997	225	152146	80.341	ng	98
35) Caprolactam	11.608	113	77479	84.477	ng	# 82
36) 4-Chloro-3-methylphenol	11.937	107	262810	83.772	ng	100
37) 2-Methylnaphthalene	12.313	142	524206	78.886	ng	97
38) 1-Methylnaphthalene	12.536	142	511482	78.566	ng	98
40) 1,2,4,5-Tetrachloroben...	12.683	216	271079	79.289	ng	99
41) Hexachlorocyclopentadiene	12.671	237	87700	91.141	ng	95
43) 2,4,6-Trichlorophenol	12.924	196	175233	86.966	ng	97
44) 2,4,5-Trichlorophenol	12.995	196	197831	88.360	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG030525\
 Data File : BG064052.D
 Acq On : 5 Mar 2025 13:44
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC080

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 03/06/2025
 Supervised By :mohammad ahmed 03/07/2025

Quant Time: Mar 05 15:25:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG030525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 05 14:45:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	13.329	154	706656	78.106	ng	98
47) 2-Chloronaphthalene	13.371	162	528541	80.099	ng	98
48) 2-Nitroaniline	13.570	65	195363	79.407	ng	97
49) Acenaphthylene	14.217	152	822307	78.787	ng	99
50) Dimethylphthalate	13.958	163	695598	78.689	ng	100
51) 2,6-Dinitrotoluene	14.070	165	147719	79.965	ng	98
52) Acenaphthene	14.557	154	540444	76.999	ng	97
53) 3-Nitroaniline	14.399	138	154723	90.561	ng	95
54) 2,4-Dinitrophenol	14.599	184	61808	82.400	ng	# 81
55) Dibenzofuran	14.892	168	874479	77.066	ng	99
56) 4-Nitrophenol	14.698	139	126982	88.621	ng	95
57) 2,4-Dinitrotoluene	14.857	165	203301	79.564	ng	# 97
58) Fluorene	15.539	166	668072	75.593	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.116	232	188307	86.273	ng	# 95
60) Diethylphthalate	15.321	149	758650	79.055	ng	99
61) 4-Chlorophenyl-phenyle...	15.539	204	326141	74.261	ng	91
62) 4-Nitroaniline	15.562	138	159478	86.459	ng	91
63) Azobenzene	15.832	77	787412	76.893	ng	99
65) 4,6-Dinitro-2-methylph...	15.615	198	100338	81.905	ng	96
66) n-Nitrosodiphenylamine	15.750	169	590976	79.884	ng	97
67) 4-Bromophenyl-phenylether	16.432	248	222258	83.033	ng	96
68) Hexachlorobenzene	16.543	284	240907	80.390	ng	98
70) Pentachlorophenol	16.884	266	170842	91.819	ng	98
71) Phenanthrene	17.278	178	1094187	78.492	ng	99
72) Anthracene	17.366	178	1099164	79.297	ng	97
73) Carbazole	17.636	167	999549	77.234	ng	99
74) Di-n-butylphthalate	18.206	149	1263186	82.920	ng	99
75) Fluoranthene	19.287	202	1276040	75.929	ng	99
77) Benzidine	19.463	184	279417	77.211	ng	99
78) Pyrene	19.646	202	1321748	79.539	ng	99
80) Butylbenzylphthalate	20.550	149	517770	80.379	ng	99
81) Benzo(a)anthracene	21.449	228	1325828	80.289	ng	98
82) 3,3'-Dichlorobenzidine	21.367	252	423636	79.269	ng	100
83) Chrysene	21.514	228	1291089	78.391	ng	99
84) Bis(2-ethylhexyl)phtha...	21.385	149	806194	90.273	ng	99
85) Di-n-octyl phthalate	22.530	149	1379517	89.556	ng	99
87) Indeno(1,2,3-cd)pyrene	27.895	276	1530546	83.318	ng	99
88) Benzo(b)fluoranthene	23.541	252	1356996	81.758	ng	98
89) Benzo(k)fluoranthene	23.606	252	1334083	80.120	ng	99
90) Benzo(a)pyrene	24.346	252	1200152	81.191	ng	97
91) Dibenzo(a,h)anthracene	27.959	278	1273882	83.647	ng	98
92) Benzo(g,h,i)perylene	28.970	276	1282909	82.048	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

