

Surrogate Summary

SW-846

SDG No.: Q1666

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167376BL	PB167376BL	2-Fluorophenol	150	157	105		15 (10)	110 (139)
		Phenol-d6	150	155	103		15 (10)	110 (134)
		Nitrobenzene-d5	100	111	111		30 (49)	130 (133)
		2-Fluorobiphenyl	100	96.5	97		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	190	127	*	15 (44)	110 (137)
		Terphenyl-d14	100	105	105		30 (48)	130 (125)
PB167376BS	PB167376BS	2-Fluorophenol	150	149	100		15 (10)	110 (139)
		Phenol-d6	150	148	98		15 (10)	110 (134)
		Nitrobenzene-d5	100	103	103		30 (49)	130 (133)
		2-Fluorobiphenyl	100	89.2	89		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	177	118	*	15 (44)	110 (137)
		Terphenyl-d14	100	97.6	98		30 (48)	130 (125)
PB167376BSD	PB167376BSD	2-Fluorophenol	150	145	97		15 (10)	110 (139)
		Phenol-d6	150	145	97		15 (10)	110 (134)
		Nitrobenzene-d5	100	104	104		30 (49)	130 (133)
		2-Fluorobiphenyl	100	87.4	87		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	173	115	*	15 (44)	110 (137)
		Terphenyl-d14	100	98.4	98		30 (48)	130 (125)
Q1666-01	TW-WTS-05	2-Fluorophenol	150	54.9	37		15 (10)	110 (139)
		Phenol-d6	150	32.1	21		15 (10)	110 (134)
		Nitrobenzene-d5	100	104	104		30 (49)	130 (133)
		2-Fluorobiphenyl	100	92.5	93		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	162	108		15 (44)	110 (137)
		Terphenyl-d14	100	90.2	90		30 (48)	130 (125)