

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1666

Client: ENTACT

Analytical Method: 8270E DataFile: BG064170.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167376BS	Phenol	50	54.3	ug/L	109				20 (66)	160 (118)	
	1,4-Dichlorobenzene	50	47.8	ug/L	96				70 (76)	130 (103)	
	1,2,4-Trichlorobenzene	50	47.0	ug/L	94				70 (41)	130 (115)	
	Naphthalene	50	46.5	ug/L	93				70 (64)	130 (107)	