

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666 SDG No.: Q1666
 Instrument ID: MSVOA_N Calibration Date(s): 03/18/2025 03/18/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:44 14:09
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085996.D		RRF050 = VN085997.D		RRF020 = VN085998.D			
		RRF010 = VN085999.D		RRF005 = VN086000.D		RRF001 = VN086001.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
Methyl tert-butyl Ether	1.888	1.691	1.578	1.604	1.638	1.673	1.679	6.6	
Carbon Tetrachloride	0.673	0.578	0.557	0.580	0.613	0.624	0.604	6.9	
Chloroform	1.119	1.017	1.011	1.101	1.149	1.245	1.107	7.9	
1,1,1-Trichloroethane	1.093	0.986	0.962	1.025	1.097	1.124	1.048	6.3	
Benzene	1.610	1.393	1.348	1.386	1.453	1.466	1.443	6.5	
Toluene	1.074	0.915	0.862	0.878	0.856	0.727	0.885	12.7	
Tetrachloroethene	0.407	0.371	0.370	0.390	0.421	0.433	0.399	6.6	
Ethyl Benzene	2.209	1.918	1.744	1.769	1.755	1.586	1.830	11.7	
m/p-Xylenes	0.867	0.756	0.700	0.690	0.660	0.643	0.719	11.4	
o-Xylene	0.831	0.713	0.655	0.645	0.585	0.532	0.660	15.8	
1,2-Dichloroethane-d4	0.632	0.665	0.669	0.721	0.737		0.685	6.3	
Dibromofluoromethane	0.333	0.334	0.347	0.362	0.368		0.349	4.5	
Toluene-d8	1.309	1.266	1.253	1.289	1.217		1.267	2.8	
4-Bromofluorobenzene	0.514	0.466	0.447	0.440	0.391		0.452	9.8	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.