

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666 SDG No.: Q1666

Instrument ID: MSVOA_N Calibration Date/Time: 03/28/2025 10:44

Lab File ID: VN086140.D Init. Calib. Date(s): 03/18/2025 03/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:44 14:09

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Methyl tert-butyl Ether	1.679	1.680		0.06	20
Carbon Tetrachloride	0.604	0.595		-1.49	20
Chloroform	1.107	1.110		0.27	20
1,1,1-Trichloroethane	1.048	1.006		-4.01	20
Benzene	1.443	1.525		5.68	20
Toluene	0.885	1.009		14.01	20
Tetrachloroethene	0.399	0.389		-2.51	20
Ethyl Benzene	1.830	1.951		6.61	20
m/p-Xylenes	0.719	0.796		10.71	20
o-Xylene	0.660	0.739		11.97	20
1,2-Dichloroethane-d4	0.685	0.616		-10.07	20
Dibromofluoromethane	0.349	0.328		-6.02	20
Toluene-d8	1.267	1.250		-1.34	20
4-Bromofluorobenzene	0.452	0.459		1.55	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.