

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1666

Client: ENTACT

Analytical Method: SW8260-Low

Datafile : VN086143.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0328WBS01	Methyl tert-butyl Ether	20	18.4	ug/L	92			70 (78)	130 (114)	
	Carbon Tetrachloride	20	19.1	ug/L	96			70 (77)	130 (113)	
	Chloroform	20	20.0	ug/L	100			70 (79)	130 (113)	
	1,1,1-Trichloroethane	20	19.1	ug/L	96			70 (80)	130 (108)	
	Benzene	20	20.3	ug/L	102			70 (82)	130 (109)	
	Toluene	20	20.8	ug/L	104			70 (82)	130 (110)	
	Tetrachloroethene	20	20.2	ug/L	101			70 (67)	130 (123)	
	Ethyl Benzene	20	19.5	ug/L	98			70 (83)	130 (109)	
	m/p-Xylenes	40	40.6	ug/L	102			70 (82)	130 (110)	
	o-Xylene	20	20.0	ug/L	100			70 (83)	130 (109)	