

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1666
Client: ENTACT
Analytical Method: SW8260-Low **Datafile :** VN086144.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0328WBSD01	Methyl tert-butyl Ether	20	19.6	ug/L	98	6		70 (78)	130 (114)	20 (20)
	Carbon Tetrachloride	20	18.2	ug/L	91	5		70 (77)	130 (113)	20 (20)
	Chloroform	20	20.4	ug/L	102	2		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	18.7	ug/L	94	2		70 (80)	130 (108)	20 (20)
	Benzene	20	20.3	ug/L	102	0		70 (82)	130 (109)	20 (20)
	Toluene	20	20.9	ug/L	104	0		70 (82)	130 (110)	20 (20)
	Tetrachloroethene	20	19.3	ug/L	97	4		70 (67)	130 (123)	20 (20)
	Ethyl Benzene	20	18.5	ug/L	93	5		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	38.1	ug/L	95	7		70 (82)	130 (110)	20 (20)
	o-Xylene	20	19.7	ug/L	99	1		70 (83)	130 (109)	20 (20)