

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666 SDG NO.: Q1666
 Lab File ID: VN086140.D Date Analyzed: 03/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:44
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	218266	8.22	341726	9.10	321573	11.87
UPPER LIMIT	436532	8.724	683452	9.6	643146	12.365
LOWER LIMIT	109133	7.724	170863	8.6	160787	11.365
EPA SAMPLE NO.						
TW-WTS-05	186773	8.22	348886	9.10	315725	11.87
VN0328WBL01	166500	8.22	307532	9.10	270205	11.87
VN0328WBS01	239264	8.22	387346	9.10	347341	11.87
VN0328WBSD01	181135	8.22	293753	9.10	265579	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.