

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ENTA05
Lab Code: CHEM Case No.: Q1666 SAS No.: Q1666 SDG NO.: Q1666
Lab File ID: VN086139.D BFB Injection Date: 03/28/2025
Instrument ID: MSVOA_N BFB Injection Time: 10:11
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.4
75	30.0 - 60.0% of mass 95	49.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	1.4 (1.6) 1
174	50.0 - 100.0% of mass 95	88
175	5.0 - 9.0% of mass 174	6.9 (7.8) 1
176	95.0 - 101.0% of mass 174	83.6 (95) 1
177	5.0 - 9.0% of mass 176	5.9 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN086140.D	03/28/2025	10:44
VN0328WBL01	VN0328WBL01	VN086142.D	03/28/2025	11:43
VN0328WBS01	VN0328WBS01	VN086143.D	03/28/2025	12:07
VN0328WBSD01	VN0328WBSD01	VN086144.D	03/28/2025	12:41
TW-WTS-05	Q1666-01	VN086161.D	03/28/2025	19:30