

Duplicate Sample Summary

Lab Sample Id :	Q1668-01DUP	Q1668-01
Client Id :	IA-B1-4-3252025DUP	IA-B1-4-3252025
DF :	1	1
Datafile :	VL042362.D	VL042361.D
Anal Date & Time :	04/17/2025 17:51	04/17/2025 17:17

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	3.9	3.8	2.6
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	1.3	1.3	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0.17	0.19	11.1
2-Butanone	0.69	0.69	0
2-Chlorotoluene	0	0	0
4-Ethyltoluene	1.1	1	9.5
4-Methyl-2-Pentanone	19.4	19.6	1
Acetone	5.2	5.3	1.9
Allyl Chloride	0	0	0
Benzene	0.53	0.56	5.5
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0.6	0.63	4.9
Carbon Tetrachloride	0.08	0.08	0

Duplicate Sample Summary

Lab Sample Id :	Q1668-01DUP	Q1668-01
Client Id :	IA-B1-4-3252025DUP	IA-B1-4-3252025
DF :	1	1
Datafile :	VL042362.D	VL042361.D
Anal Date & Time :	04/17/2025 17:51	04/17/2025 17:17

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.47	0.53	12
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	3.4	3.3	3
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.45	0.46	2.2
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0.95	0.94	1.1
Heptane	1.9	1.9	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	2.9	3	3.4
m/p-Xylene	9.2	9.4	2.2
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	2.8	2.8	0
Naphthalene	0.63	0.6	4.9
o-Xylene	2	2	0
Styrene	1.6	1.6	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0.04	0.05	22.2 *
Tetrahydrofuran	0	0	0
Toluene	1.6	1.6	0
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.23	0.24	4.3
Vinyl Chloride	0	0	0