

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
Lab Code: CHEM Case No.: Q1668 SAS No.: Q1668 SDG NO.: Q1668
Lab File ID: VL042349.D Date Analyzed: 04/17/2025
Instrument ID: MSVOA_L Time Analyzed: 09:44
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	125455	2.79	366544	3.96	338126	8.88
UPPER LIMIT	175637	3.12	513162	4.29	473376	9.21
LOWER LIMIT	75273	2.46	219926	3.63	202876	8.55
EPA SAMPLE NO.						
IA-B1-4-3252025	118680	2.79	339191	3.96	303501	8.89
IA-B1-4-3252025DL	104470	2.79	294101	3.96	258444	8.88
IA-B1-4-3252025DUP	122226	2.79	344143	3.96	303859	8.88
IA-B1-2-3252025	121182	2.79	339124	3.96	302443	8.88
IA-B1-3-3252025	117934	2.79	342044	3.96	304398	8.88
IA-B1-3-3252025DL	102685	2.79	291278	3.96	253844	8.88
IA-B1-1-3252025	117691	2.79	343133	3.96	303219	8.89
VL0417ABL01	113566	2.79	328716	3.97	297633	8.89
VL0417ABS01	122375	2.79	352201	3.97	321521	8.89

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.