

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WOOD06

Lab Code: CHEM Case No.: Q1669 SAS No.: Q1669 SDG No.: Q1669

Instrument ID: MSVOA_L Calibration Date/Time: 03/31/2025 08:51

Lab File ID: VL042201.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.476	1.509		2.24	30
Chloromethane	0.603	0.570		-5.47	30
Vinyl Chloride	0.639	0.550		-13.93	30
Bromomethane	0.273	0.252		-7.69	30
Chloroethane	0.242	0.221		-8.68	30
Tetrahydrofuran	0.418	0.371		-11.24	30
Trichlorofluoromethane	1.437	1.435		-0.14	30
1,1,2-Trichlorotrifluoroethane	1.181	1.132		-4.15	30
Dichlorotetrafluoroethane	1.257	1.261		0.32	30
tert-Butyl alcohol	1.742	1.563		-10.28	30
Heptane	1.839	1.622		-11.8	30
1,1-Dichloroethene	0.552	0.513		-7.07	30
Acetone	1.399	1.200		-14.22	30
Carbon Disulfide	1.503	1.425		-5.19	30
Methyl tert-Butyl Ether	0.873	1.103		26.35	30
Methylene Chloride	0.521	0.420		-19.39	30
trans-1,2-Dichloroethene	0.638	0.569		-10.81	30
1,1-Dichloroethane	1.183	1.151		-2.7	30
Cyclohexane	1.217	1.115		-8.38	30
2-Butanone	0.725	0.655		-9.65	30
Carbon Tetrachloride	0.781	0.807		3.33	30
cis-1,2-Dichloroethene	1.423	1.359		-4.5	30
Chloroform	2.052	2.093		2	30
1,1,1-Trichloroethane	2.349	2.228		-5.15	30
2,2,4-Trimethylpentane	1.675	1.532		-8.54	30
Benzene	0.991	0.927		-6.46	30
1,2-Dichloroethane	0.568	0.593		4.4	30
Trichloroethene	0.414	0.391		-5.56	30
1,2-Dichloropropane	0.361	0.333		-7.76	30
Bromodichloromethane	0.782	0.811		3.71	30
4-Methyl-2-Pentanone	1.149	0.949		-17.41	30
Toluene	1.175	1.070		-8.94	30
t-1,3-Dichloropropene	0.450	0.413		-8.22	30
cis-1,3-Dichloropropene	0.552	0.505		-8.51	30
1,1,2-Trichloroethane	0.382	0.366		-4.19	30
Dibromochloromethane	0.647	0.659		1.86	30
1,2-Dibromoethane	0.532	0.520		-2.26	30
Tetrachloroethene	0.393	0.337		-14.25	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WOOD06

Lab Code: CHEM Case No.: Q1669 SAS No.: Q1669 SDG No.: Q1669

Instrument ID: MSVOA_L Calibration Date/Time: 03/31/2025 08:51

Lab File ID: VL042201.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Chlorobenzene	0.949	0.917		-3.37	30
Ethyl Benzene	1.776	1.678		-5.52	30
m/p-Xylene	1.418	1.371		-3.32	30
o-Xylene	1.422	1.355		-4.71	30
Styrene	0.703	0.663		-5.69	30
Bromoform	0.581	0.568		-2.24	30
1,1,2,2-Tetrachloroethane	0.864	0.822		-4.86	30
2-Chlorotoluene	1.650	1.593		-3.45	30
1,3,5-Trimethylbenzene	1.546	1.501		-2.91	30
1,2,4-Trimethylbenzene	1.781	1.639		-7.97	30
1,3-Dichlorobenzene	0.977	0.967		-1.02	30
1,4-Dichlorobenzene	0.972	0.974		0.21	30
1,2-Dichlorobenzene	0.951	0.940		-1.16	30
1,2,4-Trichlorobenzene	0.899	0.879		-2.22	30
Hexachloro-1,3-Butadiene	0.968	0.813		-16.01	30
1,3-Butadiene	0.683	0.595		-12.88	30
Naphthalene	1.948	1.962		0.72	30
4-Ethyltoluene	1.770	1.731		-2.2	30
1-Bromo-4-Fluorobenzene	0.821	0.837		1.95	30
Hexane	1.402	1.295		-7.63	30
Allyl Chloride	0.986	0.895		-9.23	30
1,4-Dioxane	191.626	167.508		-12.59	30
Methyl Methacrylate	0.420	0.370		-11.9	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.