

Duplicate Sample Summary

Lab Sample Id :	Q1669-01DUP	Q1669-01
Client Id :	IA-B9-3-3262025DUP	IA-B9-3-3262025
DF :	1	1
Datafile :	VL042196.D	VL042185.D
Anal Date & Time :	03/28/2025 00:44	03/27/2025 18:22

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0.49	0.51	4
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0.16	0	200 *
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0.26	0.27	3.8
2-Butanone	0.93	0.91	2.2
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0.35	0.36	2.8
Acetone	36.2	34.7	4.2
Allyl Chloride	0	0	0
Benzene	0.62	0.64	3.2
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.07	0.06	15.4

Duplicate Sample Summary

Lab Sample Id :	Q1669-01DUP	Q1669-01
Client Id :	IA-B9-3-3262025DUP	IA-B9-3-3262025
DF :	1	1
Datafile :	VL042196.D	VL042185.D
Anal Date & Time :	03/28/2025 00:44	03/27/2025 18:22

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.37	0.37	0
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0.18	0	200 *
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.42	0.4	4.9
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0.28	0.31	10.2
Heptane	0.36	0.38	5.4
Hexachloro-1,3-Butadiene	0	0	0
Hexane	1.8	1.9	5.4
m/p-Xylene	1.1	1.1	0
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	0.36	0.36	0
Naphthalene	0	0	0
o-Xylene	0.49	0.51	4
Styrene	0.11	0.12	8.7
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0.52	0.5	3.9
Tetrachloroethene	0	0	0
Tetrahydrofuran	0	0	0
Toluene	2.3	2.3	0
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.22	0.22	0
Vinyl Chloride	0	0	0

Duplicate Sample Summary

Lab Sample Id :	Q1669-06DUP	Q1669-06
Client Id :	SSSG-DUP-1-3262025DUF	SSSG-DUP-1-3262025
DF :	1	1
Datafile :	VL042206.D	VL042204.D
Anal Date & Time :	03/31/2025 12:26	03/31/2025 11:06

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	1.3	1.4	7.4
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0.43	0.43	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	2.3	2.3	0
2-Butanone	6.4	6.2	3.2
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0.43	0.44	2.3
4-Methyl-2-Pentanone	0.41	0.4	2.5
Acetone	22.2	24	7.8
Allyl Chloride	0	0	0
Benzene	5.8	5.6	3.5
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.07	0.07	0

Duplicate Sample Summary

Lab Sample Id :	Q1669-06DUP	Q1669-06
Client Id :	SSSG-DUP-1-3262025DUF	SSSG-DUP-1-3262025
DF :	1	1
Datafile :	VL042206.D	VL042204.D
Anal Date & Time :	03/31/2025 12:26	03/31/2025 11:06

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.42	0.43	2.4
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	1.5	1.5	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.45	0.46	2.2
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	1.4	1.3	7.4
Heptane	1.7	1.6	6.1
Hexachloro-1,3-Butadiene	0	0	0
Hexane	4.6	4.4	4.4
m/p-Xylene	5.1	5.1	0
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	0.53	0.51	3.8
Naphthalene	0.37	0.43	15
o-Xylene	1.9	1.9	0
Styrene	1.4	1.4	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0.04	0	200 *
Tetrahydrofuran	1.9	2	5.1
Toluene	10	9.7	3
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.19	0.2	5.1
Vinyl Chloride	0	0	0

Duplicate Sample Summary

Lab Sample Id :	Q1669-09DUP	Q1669-09
Client Id :	AA-3-3252025DUP	AA-3-3252025
DF :	1	1
Datafile :	VL042190.D	VL042189.D
Anal Date & Time :	03/27/2025 21:24	03/27/2025 20:48

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0.29	0.29	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	0.23	0.18	24.4 *
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	2.3	2.5	8.3
Allyl Chloride	0	0	0
Benzene	0.15	0.13	14.3
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.07	0.07	0

Duplicate Sample Summary

Lab Sample Id :	Q1669-09DUP	Q1669-09
Client Id :	AA-3-3252025DUP	AA-3-3252025
DF :	1	1
Datafile :	VL042190.D	VL042189.D
Anal Date & Time :	03/27/2025 21:24	03/27/2025 20:48

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.45	0.44	2.2
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.43	0.45	4.5
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0	0	0
Heptane	0	0	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	3.9	4.5	14.3
m/p-Xylene	0.13	0	200 *
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	6.4	7	9
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0	0	0
Tetrahydrofuran	0	0	0
Toluene	0.33	0.34	3
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.2	0.22	9.5
Vinyl Chloride	0	0	0