

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
 Lab Code: CHEM Case No.: Q1669 SAS No.: Q1669 SDG NO.: Q1669
 Lab File ID: VL042172.D Date Analyzed: 03/27/2025
 Instrument ID: MSVOA_L Time Analyzed: 08:26
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	187002	2.78	545068	3.95	519536	8.88
UPPER LIMIT	261803	3.11	763095	4.28	727350	9.21
LOWER LIMIT	112201	2.45	327041	3.62	311722	8.55
EPA SAMPLE NO.						
IA-B9-3-3262025	182164	2.79	537249	3.96	488802	8.89
IA-B9-3-3262025DL	175494	2.78	495493	3.96	449375	8.88
IA-B9-3-3262025DUP	169076	2.78	480272	3.96	439128	8.88
IA-B9-2-3262025	187729	2.78	538422	3.96	491411	8.88
IA-B9-1-3262025	185410	2.78	523694	3.96	484321	8.88
AA-2-3252025	185847	2.79	529516	3.96	474604	8.88
SSSG-B16-3-3262025	172645	2.79	493948	3.97	445312	8.89
SSSG-B16-1-3262025	175214	2.79	504192	3.96	456695	8.88
SSSG-B16-1-3262025DL	182834	2.79	517445	3.96	472334	8.88
SSSG-B16-2-3262025	181183	2.79	509062	3.96	458107	8.88
AA-3-3252025	185013	2.79	523099	3.96	477151	8.88
AA-3-3252025DUP	184637	2.79	532429	3.96	472897	8.88
VL0327ABL01	170921	2.79	483563	3.96	446480	8.88
VL0327ABS01	166105	2.79	483209	3.96	452250	8.88

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.