

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WOOD06
Lab Code: CHEM Case No.: Q1669 SAS No.: Q1669 SDG NO.: Q1669
Lab File ID: VL042201.D Date Analyzed: 03/31/2025
Instrument ID: MSVOA_L Time Analyzed: 08:51
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	161478	2.80	460403	3.98	424671	8.90
UPPER LIMIT	226069	3.13	644564	4.31	594539	9.23
LOWER LIMIT	96886.8	2.47	276242	3.65	254803	8.57
EPA SAMPLE NO.						
SSSG-DUP-1-3262025	184538	2.81	528809	3.98	477707	8.91
SSSG-DUP-1-3262025DL	188934	2.80	537109	3.99	480959	8.91
SSSG-DUP-1-3262025DUP	184777	2.81	533717	3.99	486289	8.91
VL0331ABL01	181807	2.80	517775	3.98	464953	8.90
VL0331ABS01	175685	2.80	498986	3.98	456400	8.90

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.